

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062121\
 Data File : VW019240.D
 Acq On : 21 Jun 2021 09:49
 Operator : SY/VA
 Sample : VSTDIC005
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 21 11:30:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 21 11:29:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	365123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	622334	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	558898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	264064	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	22082	4.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	9.72%	
35) Dibromofluoromethane	7.891	113	20212	4.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	9.68%	
50) Toluene-d8	10.323	98	75819	4.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	8.88%	
62) 4-Bromofluorobenzene	12.621	95	26364	4.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	8.80%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.014	85	13496	5.28	ug/l	99
3) Chloromethane	2.215	50	17428	6.08	ug/l	96
4) Vinyl Chloride	2.367	62	25703	5.78	ug/l	97
5) Bromomethane	2.782	94	19357	5.44	ug/l	99
6) Chloroethane	2.928	64	15633	5.31	ug/l	98
7) Trichlorofluoromethane	3.269	101	12482	4.86	ug/l	98
8) Diethyl Ether	3.690	74	10201	5.08	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.074	101	19574	5.07	ug/l	95
10) Methyl Iodide	4.288	142	26226	4.83	ug/l	100
11) Tert butyl alcohol	5.190	59	7360	33.16	ug/l #	82
12) 1,1-Dichloroethene	4.050	96	19769	4.93	ug/l	85
13) Acrolein	3.903	56	5499	20.18	ug/l	93
14) Allyl chloride	4.678	41	26938	5.03	ug/l	98
15) Acrylonitrile	5.379	53	20055	26.66	ug/l	99
16) Acetone	4.135	43	20263	28.81	ug/l	98
17) Carbon Disulfide	4.397	76	59230	5.37	ug/l	97
18) Methyl Acetate	4.678	43	9361	4.99	ug/l	97
19) Methyl tert-butyl Ether	5.434	73	25708	5.19	ug/l	98
20) Methylene Chloride	4.928	84	44234	8.69	ug/l	95
21) trans-1,2-Dichloroethene	5.434	96	22557	5.05	ug/l	97
22) Diisopropyl ether	6.318	45	56054	5.31	ug/l	95
23) Vinyl Acetate	6.263	43	157440	22.96	ug/l	98
24) 1,1-Dichloroethane	6.226	63	41825	5.18	ug/l	98
25) 2-Butanone	7.183	43	25683	24.87	ug/l	98
26) 2,2-Dichloropropane	7.171	77	24688	5.40	ug/l	94
27) cis-1,2-Dichloroethene	7.171	96	23438	4.77	ug/l	92
28) Bromochloromethane	7.513	49	15989	4.96	ug/l	89
29) Tetrahydrofuran	7.543	42	15252	25.79	ug/l	94
30) Chloroform	7.683	83	41890	5.16	ug/l	95
31) Cyclohexane	7.958	56	43467	6.21	ug/l #	82
32) 1,1,1-Trichloroethane	7.878	97	31243	5.01	ug/l	98
36) 1,1-Dichloropropene	8.086	75	32658	4.77	ug/l	99
37) Ethyl Acetate	7.263	43	11529	4.68	ug/l	98
38) Carbon Tetrachloride	8.073	117	28198	4.67	ug/l	98
39) Methylcyclohexane	9.335	83	35912	4.66	ug/l	98
40) Benzene	8.329	78	90240	4.84	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062121\
 Data File : VW019240.D
 Acq On : 21 Jun 2021 09:49
 Operator : SY/VA
 Sample : VSTDIC005
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 21 11:30:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 21 11:29:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.488	41	5391	3.75	ug/l	93
42) 1,2-Dichloroethane	8.403	62	26980	4.91	ug/l	100
43) Isopropyl Acetate	8.427	43	21423	4.52	ug/l	99
44) Trichloroethene	9.098	130	22175	4.57	ug/l	86
45) 1,2-Dichloropropane	9.372	63	22304	4.92	ug/l	92
46) Dibromomethane	9.463	93	11131	4.74	ug/l	98
47) Bromodichloromethane	9.646	83	28059	4.63	ug/l	97
48) Methyl methacrylate	9.439	41	8664	3.88	ug/l	95
49) 1,4-Dioxane	9.457	88	2753	96.22	ug/l	87
51) 4-Methyl-2-Pentanone	10.213	43	54243	22.49	ug/l	98
52) Toluene	10.390	92	52879	4.46	ug/l	95
53) t-1,3-Dichloropropene	10.610	75	26293	4.37	ug/l	99
54) cis-1,3-Dichloropropene	10.079	75	30585	4.29	ug/l	95
55) 1,1,2-Trichloroethane	10.792	97	15850	4.78	ug/l	94
56) Ethyl methacrylate	10.646	69	17012	4.10	ug/l	95
57) 1,3-Dichloropropane	10.933	76	28598	4.82	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.927	63	42206	20.13	ug/l	96
59) 2-Hexanone	10.969	43	35266	21.72	ug/l	94
60) Dibromochloromethane	11.128	129	15969	4.33	ug/l	97
61) 1,2-Dibromoethane	11.238	107	14095	4.53	ug/l	100
64) Tetrachloroethene	10.866	164	17928	4.60	ug/l	92
65) Chlorobenzene	11.658	112	56733	4.66	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.731	131	18687	4.52	ug/l	99
67) Ethyl Benzene	11.731	91	103730	4.50	ug/l	100
68) m/p-Xylenes	11.841	106	74549	8.81	ug/l	95
69) o-Xylene	12.164	106	34013	4.24	ug/l	94
70) Styrene	12.182	104	55155	4.14	ug/l	98
71) Bromoform	12.347	173	8603	4.30	ug/l #	93
73) Isopropylbenzene	12.463	105	92062	4.35	ug/l	100
74) N-amyl acetate	12.274	43	18240	4.31	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.713	83	17937	4.97	ug/l	98
76) 1,2,3-Trichloropropane	12.768	75	21791	8.77	ug/l #	100
77) Bromobenzene	12.749	156	20581	4.55	ug/l	92
78) n-propylbenzene	12.804	91	119149	4.54	ug/l	98
79) 2-Chlorotoluene	12.890	91	68892	4.64	ug/l	97
80) 1,3,5-Trimethylbenzene	12.945	105	77056	4.33	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.518	75	4501	4.11	ug/l	93
82) 4-Chlorotoluene	12.987	91	72158	4.62	ug/l	98
83) tert-Butylbenzene	13.207	119	64849	4.27	ug/l	98
84) 1,2,4-Trimethylbenzene	13.249	105	78388	4.41	ug/l	95
85) sec-Butylbenzene	13.383	105	106704	4.66	ug/l	96
86) p-Isopropyltoluene	13.499	119	82563	4.28	ug/l	98
87) 1,3-Dichlorobenzene	13.499	146	45354	4.84	ug/l	98
88) 1,4-Dichlorobenzene	13.579	146	44657	4.84	ug/l	93
89) n-Butylbenzene	13.822	91	86329	4.46	ug/l	98
90) Hexachloroethane	14.091	117	15717	5.07	ug/l	88
91) 1,2-Dichlorobenzene	13.871	146	38803	4.77	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.481	75	2857	5.00	ug/l	88
93) 1,2,4-Trichlorobenzene	15.133	180	23889	4.26	ug/l	98
94) Hexachlorobutadiene	15.237	225	14075	4.44	ug/l	96
95) Naphthalene	15.365	128	38014	3.87	ug/l	98
96) 1,2,3-Trichlorobenzene	15.554	180	20876	4.40	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062121\
Data File : VW019240.D
Acq On : 21 Jun 2021 09:49
Operator : SY/VA
Sample : VSTDIC005
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :


Quant Time: Jun 21 11:30:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 21 11:29:33 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 21 11:30:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 21 11:29:33 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.952 min Scan# 992
Delta R.T. 0.001 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:168 Resp: 365123
Ion Ratio Lower Upper
168 100
99 62.2 48.8 73.2



#2
Dichlorodifluoromethane
Concen: 5.28 ug/l
RT: 2.014 min Scan# 18
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 85 Resp: 13496
Ion Ratio Lower Upper
85 100
87 33.0 16.7 50.0









#3
Chloromethane
Concen: 6.08 ug/l
RT: 2.215 min Scan# 51
Delta R.T. -0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 50 Resp: 17428
Ion Ratio Lower Upper
50 100
52 31.2 26.9 40.3



#4
Vinyl Chloride
Concen: 5.78 ug/l
RT: 2.367 min Scan# 76
Delta R.T. -0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 62 Resp: 25703
Ion Ratio Lower Upper
62 100
64 29.7 25.3 37.9









#5
Bromomethane
Concen: 5.44 ug/l
RT: 2.782 min Scan# 144
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 94 Resp: 19357
Ion Ratio Lower Upper
94 100
96 96.3 76.4 114.6



#6
Chloroethane
Concen: 5.31 ug/l
RT: 2.928 min Scan# 168
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 64 Resp: 15633
Ion Ratio Lower Upper
64 100
66 31.1 25.9 38.9









#7
 Trichlorofluoromethane
 Concen: 4.86 ug/l
 RT: 3.269 min Scan# 224
 Delta R.T. 0.006 min
 Lab File: VW019240.D
 Acq: 21 Jun 2021 09:49




Instrument :
 MSVOA_W
 ClientSampled :


Tgt Ion:101 Resp: 12482
 Ion Ratio Lower Upper
 101 100
 103 66.0 51.4 77.2



#8
 Diethyl Ether
 Concen: 5.08 ug/l
 RT: 3.690 min Scan# 293
 Delta R.T. 0.000 min
 Lab File: VW019240.D
 Acq: 21 Jun 2021 09:49

Tgt Ion: 74 Resp: 10201
 Ion Ratio Lower Upper
 74 100
 45 91.2 43.3 129.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.07 ug/l
RT: 4.074 min Scan# 356
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:101 Resp: 19574
Ion Ratio Lower Upper
101 100
85 46.8 35.9 53.9
151 69.1 59.8 89.6



#10
Methyl Iodide
Concen: 4.83 ug/l
RT: 4.288 min Scan# 391
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:142 Resp: 26226
Ion Ratio Lower Upper
142 100
127 42.5 33.7 50.5
141 14.1 11.2 16.8









#11
Tert butyl alcohol
Concen: 33.16 ug/l
RT: 5.190 min Scan# 539
Delta R.T. 0.018 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 59 Resp: 7360
Ion Ratio Lower Upper
59 100
57 4.3 8.9 13.3#



#12
1,1-Dichloroethene
Concen: 4.93 ug/l
RT: 4.050 min Scan# 352
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 96 Resp: 19769
Ion Ratio Lower Upper
96 100
61 176.0 123.0 184.4
98 65.3 47.5 71.3









#13
Acrolein
Concen: 20.18 ug/l
RT: 3.903 min Scan# 328
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 56 Resp: 5499
Ion Ratio Lower Upper
56 100
55 76.7 56.7 85.1



#14
Allyl chloride
Concen: 5.03 ug/l
RT: 4.678 min Scan# 455
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 41 Resp: 26938
Ion Ratio Lower Upper
41 100
39 79.5 63.9 95.9
76 40.0 35.1 52.7









#15
Acrylonitrile
Concen: 26.66 ug/l
RT: 5.379 min Scan# 570
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:	53	Resp:	20055
Ion	Ratio	Lower	Upper
53	100		
52	83.9	66.5	99.7
51	37.6	29.8	44.8



#16
Acetone
Concen: 28.81 ug/l
RT: 4.135 min Scan# 366
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	43	Resp:	20263
Ion	Ratio	Lower	Upper
43	100		
58	31.8	26.4	39.6









#17
Carbon Disulfide
Concen: 5.37 ug/l
RT: 4.397 min Scan# 409
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 76 Resp: 59230
Ion Ratio Lower Upper
76 100
78 10.3 7.4 11.2



#18
Methyl Acetate
Concen: 4.99 ug/l
RT: 4.678 min Scan# 455
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 43 Resp: 9361
Ion Ratio Lower Upper
43 100
74 25.6 21.8 32.8









#19
Methyl tert-butyl Ether
Concen: 5.19 ug/l
RT: 5.434 min Scan# 579
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 73 Resp: 25708
Ion Ratio Lower Upper
73 100
57 20.8 17.4 26.0



#20
Methylene Chloride
Concen: 8.69 ug/l
RT: 4.928 min Scan# 496
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49


Tgt Ion: 84 Resp: 44234
Ion Ratio Lower Upper
84 100
49 120.0 90.4 135.6
51 36.5 27.1 40.7
86 57.5 47.6 71.4








#21
trans-1,2-Dichloroethene
Concen: 5.05 ug/l
RT: 5.434 min Scan# 579
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Instrument :
MSVOA_W
ClientSampled :




Tgt Ion: 96 Resp: 22557

Ion	Ratio	Lower	Upper
96	100		
61	139.3	108.5	162.7
98	63.2	49.4	74.0






#22
Diisopropyl ether
Concen: 5.31 ug/l
RT: 6.318 min Scan# 724
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49



Tgt Ion: 45 Resp: 56054

Ion	Ratio	Lower	Upper
45	100		
43	61.8	46.1	69.1
87	25.9	23.4	35.2
59	11.8	9.7	14.5









#23
 Vinyl Acetate
 Concen: 22.96 ug/l
 RT: 6.263 min Scan# 715
 Delta R.T. 0.006 min
 Lab File: VW019240.D
 Acq: 21 Jun 2021 09:49




Instrument :
 MSVOA_W
 ClientSampled :


Tgt Ion: 43 Resp: 157440
 Ion Ratio Lower Upper
 43 100
 86 11.4 9.6 14.4



#24
 1,1-Dichloroethane
 Concen: 5.18 ug/l
 RT: 6.226 min Scan# 709
 Delta R.T. 0.006 min
 Lab File: VW019240.D
 Acq: 21 Jun 2021 09:49

Tgt Ion: 63 Resp: 41825
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.6 10.9
 100 4.1 2.3 6.8









#25
2-Butanone
Concen: 24.87 ug/l
RT: 7.183 min Scan# 866
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 25683
Ion Ratio Lower Upper
43 100
72 30.9 24.0 36.0



#26
2,2-Dichloropropane
Concen: 5.40 ug/l
RT: 7.171 min Scan# 864
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 77 Resp: 24688
Ion Ratio Lower Upper
77 100
97 20.3 11.8 35.3









#27
cis-1,2-Dichloroethene
Concen: 4.77 ug/l
RT: 7.171 min Scan# 864
Delta R.T. -0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 96 Resp: 23438
Ion Ratio Lower Upper
96 100
61 154.2 0.0 281.2
98 64.2 0.0 127.4



#28
Bromochloromethane
Concen: 4.96 ug/l
RT: 7.513 min Scan# 920
Delta R.T. -0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 49 Resp: 15989
Ion Ratio Lower Upper
49 100
129 2.1 0.0 4.4
130 69.2 63.0 94.4









#29
Tetrahydrofuran
Concen: 25.79 ug/l
RT: 7.543 min Scan# 925
Delta R.T. 0.012 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 42 Resp: 15252
Ion Ratio Lower Upper
42 100
72 47.1 40.1 60.1
71 43.3 38.3 57.5



#30
Chloroform
Concen: 5.16 ug/l
RT: 7.683 min Scan# 948
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 83 Resp: 41890
Ion Ratio Lower Upper
83 100
85 62.4 53.4 80.0









#31
Cyclohexane
Concen: 6.21 ug/l
RT: 7.958 min Scan# 993
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 56 Resp: 43467
Ion Ratio Lower Upper
56 100
69 60.4 28.4 42.6#
84 97.9 72.3 108.5



#32
1,1,1-Trichloroethane
Concen: 5.01 ug/l
RT: 7.878 min Scan# 980
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 97 Resp: 31243
Ion Ratio Lower Upper
97 100
99 65.2 52.2 78.4
61 48.3 36.7 55.1









#33
1,2-Dichloroethane-d4
Concen: 4.86 ug/l
RT: 8.311 min Scan# 1051
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 65 Resp: 22082
Ion Ratio Lower Upper
65 100
67 53.5 0.0 104.0



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.848 min Scan# 1139
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 114 Resp: 622334
Ion Ratio Lower Upper
114 100
63 21.3 0.0 42.4
88 15.8 0.0 33.2









#35
Dibromofluoromethane
Concen: 4.84 ug/l
RT: 7.891 min Scan# 982
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:113 Resp: 20212
Ion Ratio Lower Upper
113 100
111 100.1 82.3 123.5
192 17.8 15.3 22.9



#36
1,1-Dichloropropene
Concen: 4.77 ug/l
RT: 8.086 min Scan# 1014
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 75 Resp: 32658
Ion Ratio Lower Upper
75 100
110 32.9 17.1 51.2
77 31.4 25.3 37.9





#37

Ethyl Acetate

Concen: 4.68 ug/l

RT: 7.263 min Scan# 879





Delta R.T. 0.006 min

Lab File: VW019240.D

Acq: 21 Jun 2021 09:49

Instrument :
MSVOA_W
ClientSampled :

Tgt Ion: 43	Resp: 11529
Ion Ratio	Lower Upper
43 100	
61 16.3	12.2 18.4
70 12.7	9.5 14.3



#38

Carbon Tetrachloride

Concen: 4.67 ug/l




RT: 8.073 min Scan# 1012


Delta R.T. 0.000 min

Lab File: VW019240.D

Acq: 21 Jun 2021 09:49

Tgt Ion: 117	Resp: 28198
Ion Ratio	Lower Upper
117 100	
119 97.1	77.4 116.0
121 32.3	23.6 35.4









#39
Methylcyclohexane
Concen: 4.66 ug/l
RT: 9.335 min Scan# 1219
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 83 Resp: 35912
Ion Ratio Lower Upper
83 100
55 78.7 60.9 91.3
98 48.9 38.4 57.6



#40
Benzene
Concen: 4.84 ug/l
RT: 8.329 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 78 Resp: 90240
Ion Ratio Lower Upper
78 100
77 24.9 19.6 29.4









#41
Methacrylonitrile
Concen: 3.75 ug/l
RT: 7.488 min Scan# 916
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:	41	Resp:	5391
Ion	Ratio	Lower	Upper
41	100		
39	66.4	50.1	75.1
67	92.7	69.0	103.4
52	43.6	29.3	43.9



#42
1,2-Dichloroethane
Concen: 4.91 ug/l
RT: 8.403 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	62	Resp:	26980
Ion	Ratio	Lower	Upper
62	100		
98	8.8	0.0	18.0









#43
Isopropyl Acetate
Concen: 4.52 ug/l
RT: 8.427 min Scan# 1070
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:	43	Resp:	21423
Ion	Ratio	Lower	Upper
43	100		
61	33.1	27.0	40.6
87	13.9	11.6	17.4



#44
Trichloroethene
Concen: 4.57 ug/l
RT: 9.098 min Scan# 1180
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	130	Resp:	22175
Ion	Ratio	Lower	Upper
130	100		
95	113.4	0.0	199.0









#45
1,2-Dichloropropane
Concen: 4.92 ug/l
RT: 9.372 min Scan# 1225
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 63 Resp: 22304
Ion Ratio Lower Upper
63 100
65 34.4 24.1 36.1



#46
Dibromomethane
Concen: 4.74 ug/l
RT: 9.463 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 93 Resp: 11131
Ion Ratio Lower Upper
93 100
95 82.9 66.7 100.1
174 90.6 75.3 112.9









#47
Bromodichloromethane
Concen: 4.63 ug/l
RT: 9.646 min Scan# 1270
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:	83	Resp:	28059
Ion	Ratio	Lower	Upper
83	100		
85	64.2	53.4	80.2
127	7.4	7.0	10.6



#48
Methyl methacrylate
Concen: 3.88 ug/l
RT: 9.439 min Scan# 1236
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	41	Resp:	8664
Ion	Ratio	Lower	Upper
41	100		
69	98.5	80.2	120.2
39	73.3	51.5	77.3









#49
1,4-Dioxane
Concen: 96.22 ug/l
RT: 9.457 min Scan# 1239
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 88 Resp: 2753
Ion Ratio Lower Upper
88 100
43 30.6 22.1 33.1
58 82.7 55.4 83.2



#50
Toluene-d8
Concen: 4.44 ug/l
RT: 10.323 min Scan# 1381
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 98 Resp: 75819
Ion Ratio Lower Upper
98 100
100 64.6 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 22.49 ug/l
RT: 10.213 min Scan# 1363
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 54243
Ion Ratio Lower Upper
43 100
58 41.3 34.2 51.4



#52
Toluene
Concen: 4.46 ug/l
RT: 10.390 min Scan# 1392
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 92 Resp: 52879
Ion Ratio Lower Upper
92 100
91 177.9 137.4 206.0









#53
t-1,3-Dichloropropene
Concen: 4.37 ug/l
RT: 10.610 min Scan# 1428
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 75 Resp: 26293
Ion Ratio Lower Upper
75 100
77 30.5 24.8 37.2



#54
cis-1,3-Dichloropropene
Concen: 4.29 ug/l
RT: 10.079 min Scan# 1341
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 75 Resp: 30585
Ion Ratio Lower Upper
75 100
77 33.1 25.0 37.6
39 50.2 36.9 55.3









#55
1,1,2-Trichloroethane
Concen: 4.78 ug/l
RT: 10.792 min Scan# 1458
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 97	Resp: 15850
Ion Ratio	Lower Upper
97 100	
83 82.6	70.6 106.0
85 52.8	45.8 68.6
99 61.1	51.9 77.9



#56
Ethyl methacrylate
Concen: 4.10 ug/l
RT: 10.646 min Scan# 1434
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 69	Resp: 17012
Ion Ratio	Lower Upper
69 100	
41 63.2	48.7 73.1
39 31.6	29.0 43.4









#57
1,3-Dichloropropane
Concen: 4.82 ug/l
RT: 10.933 min Scan# 1481
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 76 Resp: 28598
Ion Ratio Lower Upper
76 100
78 31.4 25.5 38.3



#58
2-Chloroethyl Vinyl ether
Concen: 20.13 ug/l
RT: 9.927 min Scan# 1316
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 63 Resp: 42206
Ion Ratio Lower Upper
63 100
106 25.7 22.3 33.5









#59
2-Hexanone
Concen: 21.72 ug/l
RT: 10.969 min Scan# 1487
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 35266
Ion Ratio Lower Upper
43 100
58 54.0 29.0 87.1



#60
Dibromochloromethane
Concen: 4.33 ug/l
RT: 11.128 min Scan# 1513
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:129 Resp: 15969
Ion Ratio Lower Upper
129 100
127 78.7 38.2 114.6









#61
1,2-Dibromoethane
Concen: 4.53 ug/l
RT: 11.238 min Scan# 1531
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:107 Resp: 14095
Ion Ratio Lower Upper
107 100
109 92.9 73.9 110.9



#62
4-Bromofluorobenzene
Concen: 4.40 ug/l
RT: 12.621 min Scan# 1758
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 95 Resp: 26364
Ion Ratio Lower Upper
95 100
174 72.1 0.0 156.2
176 69.3 0.0 160.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.634 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:117 Resp: 558898
Ion Ratio Lower Upper
117 100
82 58.3 44.4 66.6
119 32.5 24.6 36.8



#64
Tetrachloroethene
Concen: 4.60 ug/l
RT: 10.866 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:164 Resp: 17928
Ion Ratio Lower Upper
164 100
166 123.3 107.8 161.8
129 92.6 77.3 115.9
131 88.7 78.5 117.7









#65
Chlorobenzene
Concen: 4.66 ug/l
RT: 11.658 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:112 Resp: 56733
Ion Ratio Lower Upper
112 100
114 32.4 26.4 39.6



#66
1,1,1,2-Tetrachloroethane
Concen: 4.52 ug/l
RT: 11.731 min Scan# 1612
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49


Tgt Ion:131 Resp: 18687
Ion Ratio Lower Upper
131 100
133 95.4 48.1 144.3
119 66.2 33.1 99.2








#67
Ethyl Benzene
Concen: 4.50 ug/l
RT: 11.731 min Scan# 1612
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Instrument :
MSVOA_W
ClientSampled :







Tgt Ion: 91 Resp: 103730
Ion Ratio Lower Upper
91 100
106 29.8 24.0 36.0



#68
m/p-Xylenes
Concen: 8.81 ug/l
RT: 11.841 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:106 Resp: 74549
Ion Ratio Lower Upper
106 100
91 212.5 164.2 246.2









#69
o-Xylene
Concen: 4.24 ug/l
RT: 12.164 min Scan# 1683
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:106 Resp: 34013
Ion Ratio Lower Upper
106 100
91 227.6 109.1 327.2



#70
Styrene
Concen: 4.14 ug/l
RT: 12.182 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:104 Resp: 55155
Ion Ratio Lower Upper
104 100
78 53.4 41.3 61.9
103 56.5 44.1 66.1









#71
Bromoform
Concen: 4.30 ug/l
RT: 12.347 min Scan# 1713
Delta R.T. -0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:173 Resp: 8603
Ion Ratio Lower Upper
173 100
175 44.1 24.6 73.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.560 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:152 Resp: 264064
Ion Ratio Lower Upper
152 100
115 62.2 30.3 90.8
150 161.7 0.0 356.0









#73
Isopropylbenzene
Concen: 4.35 ug/l
RT: 12.463 min Scan# 1732
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:105 Resp: 92062
Ion Ratio Lower Upper
105 100
120 26.1 13.2 39.5



#74
N-ethyl acetate
Concen: 4.31 ug/l
RT: 12.274 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 43 Resp: 18240
Ion Ratio Lower Upper
43 100
70 45.2 38.4 57.6
55 28.9 21.6 32.4
61 24.5 21.0 31.4









#75
1,1,2,2-Tetrachloroethane
Concen: 4.97 ug/l
RT: 12.713 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:	83	Resp:	17937
Ion	Ratio	Lower	Upper
83	100		
131	9.3	5.1	15.2
85	64.6	31.6	94.8



#76
1,2,3-Trichloropropane
Concen: 8.77 ug/l
RT: 12.768 min Scan# 1782
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	75	Resp:	21791
Ion	Ratio	Lower	Upper
75	100		
77	199.7	0.0	0.0#









#77
Bromobenzene
Concen: 4.55 ug/l
RT: 12.749 min Scan# 1779
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:156 Resp: 20581
Ion Ratio Lower Upper
156 100
77 211.4 96.7 290.0
158 97.7 48.4 145.3



#78
n-propylbenzene
Concen: 4.54 ug/l
RT: 12.804 min Scan# 1788
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 91 Resp: 119149
Ion Ratio Lower Upper
91 100
120 21.2 11.1 33.3









#79
2-Chlorotoluene
Concen: 4.64 ug/l
RT: 12.890 min Scan# 1802
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 91 Resp: 68892
Ion Ratio Lower Upper
91 100
126 31.4 16.5 49.5



#80
1,3,5-Trimethylbenzene
Concen: 4.33 ug/l
RT: 12.945 min Scan# 1811
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49


Tgt Ion:105 Resp: 77056
Ion Ratio Lower Upper
105 100
120 48.4 24.1 72.2








#81
trans-1,4-Dichloro-2-butene
Concen: 4.11 ug/l
RT: 12.518 min Scan# 1741
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49


Instrument :
MSVOA_W
ClientSampled :





Tgt Ion: 75 Resp: 4501
Ion Ratio Lower Upper
75 100
53 112.0 81.5 122.3
89 44.9 36.3 54.5




#82
4-Chlorotoluene
Concen: 4.62 ug/l
RT: 12.987 min Scan# 1818
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49



Tgt Ion: 91 Resp: 72158
Ion Ratio Lower Upper
91 100
126 31.0 16.1 48.2









#83
tert-Butylbenzene
Concen: 4.27 ug/l
RT: 13.207 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:119 Resp: 64849
Ion Ratio Lower Upper
119 100
91 70.7 34.0 102.0
134 25.8 13.0 38.9



#84
1,2,4-Trimethylbenzene
Concen: 4.41 ug/l
RT: 13.249 min Scan# 1861
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49


Tgt Ion:105 Resp: 78388
Ion Ratio Lower Upper
105 100
120 41.4 22.4 67.0








#85
sec-Butylbenzene
Concen: 4.66 ug/l
RT: 13.383 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Instrument :
MSVOA_W
ClientSampled :







Tgt Ion:105 Resp: 106704
Ion Ratio Lower Upper
105 100
134 18.4 10.1 30.1



#86
p-Isopropyltoluene
Concen: 4.28 ug/l
RT: 13.499 min Scan# 1902
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:119 Resp: 82563
Ion Ratio Lower Upper
119 100
134 26.0 12.8 38.3
91 26.0 12.1 36.3









#87
1,3-Dichlorobenzene
Concen: 4.84 ug/l
RT: 13.499 min Scan# 1902
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:	146	Resp:	45354
Ion	Ratio	Lower	Upper
146	100		
111	41.7	20.7	62.1
148	61.3	31.8	95.3



#88
1,4-Dichlorobenzene
Concen: 4.84 ug/l
RT: 13.579 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	146	Resp:	44657
Ion	Ratio	Lower	Upper
146	100		
111	48.7	20.1	60.3
148	66.9	32.4	97.2









#89
n-Butylbenzene
Concen: 4.46 ug/l
RT: 13.822 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:	91	Resp:	86329
Ion	Ratio	Lower	Upper
91	100		
92	52.0	26.6	79.8
134	23.8	12.7	38.0



#90
Hexachloroethane
Concen: 5.07 ug/l
RT: 14.091 min Scan# 1999
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:	117	Resp:	15717
Ion	Ratio	Lower	Upper
117	100		
201	66.7	38.4	115.2









#91
1,2-Dichlorobenzene
Concen: 4.77 ug/l
RT: 13.871 min Scan# 1963
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:146 Resp: 38803
Ion Ratio Lower Upper
146 100
111 43.9 21.1 63.1
148 61.5 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.00 ug/l
RT: 14.481 min Scan# 2063
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion: 75 Resp: 2857
Ion Ratio Lower Upper
75 100
155 71.9 41.3 124.1
157 91.1 51.2 153.6









#93
1,2,4-Trichlorobenzene
Concen: 4.26 ug/l
RT: 15.133 min Scan# 2170
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:180 Resp: 23889
Ion Ratio Lower Upper
180 100
182 96.0 48.9 146.7
145 31.4 15.4 46.1



#94
Hexachlorobutadiene
Concen: 4.44 ug/l
RT: 15.237 min Scan# 2187
Delta R.T. 0.006 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49


Tgt Ion:225 Resp: 14075
Ion Ratio Lower Upper
225 100
223 61.3 29.6 88.8
227 64.1 30.4 91.3








#95
Naphthalene
Concen: 3.87 ug/l
RT: 15.365 min Scan# 2208
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Instrument :
MSVOA_W
ClientSampleId :



Tgt Ion:128 Resp: 38014
Ion Ratio Lower Upper
128 100
127 11.7 10.4 15.6
129 11.1 8.5 12.7



#96
1,2,3-Trichlorobenzene
Concen: 4.40 ug/l
RT: 15.554 min Scan# 2239
Delta R.T. 0.000 min
Lab File: VW019240.D
Acq: 21 Jun 2021 09:49

Tgt Ion:180 Resp: 20876
Ion Ratio Lower Upper
180 100
182 93.7 47.4 142.3
145 33.4 16.1 48.2

