

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062123\
 Data File : VW026283.D
 Acq On : 21 Jun 2023 12:20
 Operator : SY/MD
 Sample : 03334-08
 Misc : 5.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YBW03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/22/2023
 Supervised By :Mahesh Dadoda 06/22/2023

Quant Time: Jun 22 01:34:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 22 01:33:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	835568	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	687283	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	239999	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	169349	14.407	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	57.640%		
7) Chloroethane-d5	2.893	69	164970	18.748	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	75.000%		
11) 1,1-Dichloroethene-d2	4.015	65	90123	15.549	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	62.200%		
21) 2-Butanone-d5	7.075	46	153325	44.803	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	89.600%		
24) Chloroform-d	7.648	84	440668	18.010	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	72.040%		
26) 1,2-Dichloroethane-d4	8.301	65	250545	18.726	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	74.920%		
32) Benzene-d6	8.270	84	841654	18.199	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	72.800%		
36) 1,2-Dichloropropane-d6	9.270	67	287099	19.596	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	78.400%		
41) Toluene-d8	10.325	98	668893	16.309	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	65.240%		
43) trans-1,3-Dichloroprop...	10.575	79	92060	15.856	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	63.440%		
47) 2-Hexanone-d5	10.922	63	87696	37.960	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	75.920%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	217582	20.408	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	81.640%		
66) 1,2-Dichlorobenzene-d4	13.848	152	171209	18.638	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	74.560%#		
Target Compounds						
5) Vinyl chloride	2.369	62	15265m	1.087	ug/L	
12) 1,1-Dichloroethene	4.039	96	321994m	30.267	ug/L	
13) Acetone	4.125	43	20952	7.912	ug/L	100
19) 1,1-Dichloroethane	6.216	63	128811	5.301	ug/L	100
33) Benzene	8.319	78	99942	2.123	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062123\
Data File : VW026283.D
Acq On : 21 Jun 2023 12:20
Operator : SY/MD
Sample : 03334-08
Misc : 5.16g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YBW03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/22/2023
Supervised By :Mahesh Dadoda 06/22/2023

Quant Time: Jun 22 01:34:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 22 01:33:30 2023
Response via : Initial Calibration

