

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062124\
 Data File : VW029297.D
 Acq On : 21 Jun 2024 12:17
 Operator : SY/MD
 Sample : P2963-02RE
 Misc : 5.31g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4B47RE

Quant Time: Jun 21 22:54:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 22:53:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	202430	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	165020	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	49187	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	52453	18.721	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.880%
7) Chloroethane-d5	2.893	69	46170	19.513	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.040%
11) 1,1-Dichloroethene-d2	4.015	65	24412	18.045	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.200%
21) 2-Butanone-d5	7.081	46	41605	49.632	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.260%
24) Chloroform-d	7.648	84	119801	19.839	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.360%
26) 1,2-Dichloroethane-d4	8.307	65	73895	21.314	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.240%
32) Benzene-d6	8.270	84	220947	22.043	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.160%
36) 1,2-Dichloropropane-d6	9.270	67	76400	23.978	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.920%
41) Toluene-d8	10.319	98	165450	18.657	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	74.640%
43) trans-1,3-Dichloroprop...	10.575	79	25643	19.972	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.880%
47) 2-Hexanone-d5	10.922	63	22697	45.442	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.880%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	57974	22.862	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	37387	21.898	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062124\
Data File : VW029297.D
Acq On : 21 Jun 2024 12:17
Operator : SY/MD
Sample : P2963-02RE
Misc : 5.31g/10mL/MSVOA_W/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4B47RE

Quant Time: Jun 21 22:54:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 22:53:01 2024
Response via : Initial Calibration

