

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062124\
 Data File : VW029309.D
 Acq On : 21 Jun 2024 17:59
 Operator : SY/MD
 Sample : P2982-10
 Misc : 5.44g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CF6

Quant Time: Jun 21 22:58:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 22:53:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	424890	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	345594	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	120562	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	107036	18.200	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.800%
7) Chloroethane-d5	2.893	69	98550	19.844	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.360%
11) 1,1-Dichloroethene-d2	4.021	65	55961	19.708	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.840%
21) 2-Butanone-d5	7.075	46	71950	40.893	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.780%
24) Chloroform-d	7.648	84	282743	22.308	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.240%
26) 1,2-Dichloroethane-d4	8.301	65	161521	22.196	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	8.270	84	541464	25.795	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.160%
36) 1,2-Dichloropropane-d6	9.270	67	180608	27.067	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.280%
41) Toluene-d8	10.319	98	448569	24.153	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.600%
43) trans-1,3-Dichloroprop...	10.575	79	55970	20.815	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.240%
47) 2-Hexanone-d5	10.922	63	50665	48.436	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.880%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	124068	23.362	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.440%
66) 1,2-Dichlorobenzene-d4	13.849	152	97751	23.358	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.440%
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	16748	2.242	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062124\
Data File : VW029309.D
Acq On : 21 Jun 2024 17:59
Operator : SY/MD
Sample : P2982-10
Misc : 5.44g/10mL/MSVOA_W/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CF6

Quant Time: Jun 21 22:58:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 22:53:01 2024
Response via : Initial Calibration

