

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062124\
 Data File : VW029312.D
 Acq On : 21 Jun 2024 19:11
 Operator : SY/MD
 Sample : P2982-13
 Misc : 5.60g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CG0

Quant Time: Jun 21 22:59:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 21 22:53:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	506579	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	421201	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	160447	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	111489	15.901	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.600%
7) Chloroethane-d5	2.899	69	99533	16.810	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.240%
11) 1,1-Dichloroethene-d2	4.021	65	53704	15.863	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.440%
21) 2-Butanone-d5	7.075	46	74552	35.539	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.080%
24) Chloroform-d	7.648	84	276943	18.327	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.320%
26) 1,2-Dichloroethane-d4	8.301	65	154301	17.784	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	71.120%
32) Benzene-d6	8.276	84	524894	20.517	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.080%
36) 1,2-Dichloropropane-d6	9.270	67	174342	21.438	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.760%
41) Toluene-d8	10.319	98	446990	19.748	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.000%
43) trans-1,3-Dichloroprop...	10.575	79	57759	17.624	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	70.480%
47) 2-Hexanone-d5	10.922	63	52408	41.108	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.220%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	122544	18.933	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	112399	20.182	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.720%
Target Compounds						
16) Methylene chloride	4.917	84	19119	2.146	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062124\
Data File : VW029312.D
Acq On : 21 Jun 2024 19:11
Operator : SY/MD
Sample : P2982-13
Misc : 5.60g/10mL/MSVOA_W/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CG0

Quant Time: Jun 21 22:59:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 21 22:53:01 2024
Response via : Initial Calibration

