

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062221\
 Data File : VW019278.D
 Acq On : 22 Jun 2021 19:22
 Operator : SY/VA
 Sample : M2812-04
 Misc : 6.12G/5ML/MSVOA_W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 23 05:44:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 21 12:12:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.952	168	37174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	60649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	49018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	17519	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	26496	58.757	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.520%	
35) Dibromofluoromethane	7.891	113	20990	54.336	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.680%	
50) Toluene-d8	10.329	98	67156	43.427	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.860%	
62) 4-Bromofluorobenzene	12.621	95	21393	39.555	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.100%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.178	59	64	Below Cal	#	71
16) Acetone	4.129	43	15257	210.801	ug/l	94
20) Methylene Chloride	4.928	84	4151	2.417	ug/l #	78
25) 2-Butanone	7.196	43	206	1.930	ug/l #	45
31) Cyclohexane	7.945	56	1281	1.671	ug/l #	63
36) 1,1-Dichloropropene	7.945	75	3395	5.289	ug/l #	48
38) Carbon Tetrachloride	7.952	117	4050	7.218	ug/l #	17
39) Methylcyclohexane	9.335	83	1013	1.331	ug/l	97
40) Benzene	8.323	78	3116	1.772	ug/l	95
52) Toluene	10.390	92	21706	19.837	ug/l	97
67) Ethyl Benzene	11.731	91	38352	20.387	ug/l	96
68) m/p-Xylenes	11.835	106	56473	80.410	ug/l	92
69) o-Xylene	12.170	106	26174	39.762	ug/l	95
70) Styrene	12.176	104	5623	5.066	ug/l #	17
73) Isopropylbenzene	12.463	105	9422	7.231	ug/l	95
74) N-amyl acetate	12.250	43	485	1.758	ug/l #	67
75) 1,1,2,2-Tetrachloroethane	12.701	83	332	1.401	ug/l #	19
76) 1,2,3-Trichloropropane	12.859	75	763	4.446	ug/l #	100
78) n-propylbenzene	12.804	91	10609	6.458	ug/l	96
79) 2-Chlorotoluene	13.109	91	4405	4.691	ug/l	92
80) 1,3,5-Trimethylbenzene	12.945	105	75607	68.440	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.615	75	13450	190.115	ug/l #	12
82) 4-Chlorotoluene	13.109	91	4405	4.488	ug/l	90
83) tert-Butylbenzene	13.249	119	19668	21.087	ug/l #	60
84) 1,2,4-Trimethylbenzene	13.249	105	173961	158.609	ug/l	99
85) sec-Butylbenzene	13.249	105	173961	119.911	ug/l #	56
86) p-Isopropyltoluene	13.499	119	8404	7.119	ug/l	90
88) 1,4-Dichlorobenzene	13.579	146	777	1.325	ug/l #	45
89) n-Butylbenzene	13.755	91	61999	52.215	ug/l #	47
90) Hexachloroethane	14.091	117	5486	25.639	ug/l #	11
92) 1,2-Dibromo-3-Chloropr...	14.481	75	2311	61.095	ug/l #	12
95) Naphthalene	15.365	128	4163280	6909.965	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062221\
Data File : VW019278.D
Acq On : 22 Jun 2021 19:22
Operator : SY/VA
Sample : M2812-04
Misc : 6.12G/5ML/MSVOA_W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :


Quant Time: Jun 23 05:44:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 21 12:12:47 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.952 min Scan# 992
Delta R.T. 0.001 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22


Instrument :
MSVOA_W
ClientSampled :





Tgt Ion:168 Resp: 37174
Ion Ratio Lower Upper
168 100
99 61.5 48.8 73.2




#11
Tert butyl alcohol
Concen: Below Cal
RT: 5.178 min Scan# 537
Delta R.T. 0.006 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22



Tgt Ion: 59 Resp: 64
Ion Ratio Lower Upper
59 100
57 0.0 8.9 13.3#









#16
Acetone
Concen: 210.801 ug/l
RT: 4.129 min Scan# 365
Delta R.T. 0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 15257
Ion Ratio Lower Upper
43 100
58 29.5 26.4 39.6



#20
Methylene Chloride
Concen: 2.417 ug/l
RT: 4.928 min Scan# 496
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 84 Resp: 4151
Ion Ratio Lower Upper
84 100
49 81.0 90.4 135.6#
51 54.8 27.1 40.7#
86 59.8 47.6 71.4









#25
2-Butanone
Concen: 1.930 ug/l
RT: 7.196 min Scan# 868
Delta R.T. 0.019 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 43 Resp: 206
Ion Ratio Lower Upper
43 100
72 0.0 24.0 36.0#



#31
Cyclohexane
Concen: 1.671 ug/l
RT: 7.945 min Scan# 991
Delta R.T. -0.013 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 56 Resp: 1281
Ion Ratio Lower Upper
56 100
69 103.3 28.4 42.6#
84 96.5 72.3 108.5









#33
 1,2-Dichloroethane-d4
 Concen: 58.757 ug/l
 RT: 8.305 min Scan# 1050
 Delta R.T. -0.006 min
 Lab File: VW019278.D
 Acq: 22 Jun 2021 19:22




Instrument :
 MSVOA_W
 ClientSampleId :


Tgt Ion: 65 Resp: 26496
 Ion Ratio Lower Upper
 65 100
 67 51.2 0.0 104.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.848 min Scan# 1139
 Delta R.T. 0.006 min
 Lab File: VW019278.D
 Acq: 22 Jun 2021 19:22

Tgt Ion: 114 Resp: 60649
 Ion Ratio Lower Upper
 114 100
 63 22.9 0.0 42.4
 88 15.7 0.0 33.2









#35
Dibromofluoromethane
Concen: 54.336 ug/l
RT: 7.891 min Scan# 982
Delta R.T. 0.007 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:113 Resp: 20990
Ion Ratio Lower Upper
113 100
111 105.7 82.3 123.5
192 19.6 15.3 22.9



#36
1,1-Dichloropropene
Concen: 5.289 ug/l
RT: 7.945 min Scan# 991
Delta R.T. -0.141 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 75 Resp: 3395
Ion Ratio Lower Upper
75 100
110 6.4 17.1 51.2#
77 0.0 25.3 37.9#





#38

Carbon Tetrachloride

Concen: 7.218 ug/l

RT: 7.952 min Scan# 992

Delta R.T. -0.121 min


Lab File: VW019278.D

Acq: 22 Jun 2021 19:22




Instrument :

MSVOA_W

ClientSampled :



Tgt Ion:	117	Resp:	4050
Ion	Ratio	Lower	Upper
117	100		
119	7.0	77.4	116.0#
121	0.0	23.6	35.4#



#39

Methylcyclohexane

Concen: 1.331 ug/l




RT: 9.335 min Scan# 1219


Delta R.T. 0.000 min

Lab File: VW019278.D

Acq: 22 Jun 2021 19:22

Tgt Ion:	83	Resp:	1013
Ion	Ratio	Lower	Upper
83	100		
55	74.4	60.9	91.3
98	51.9	38.4	57.6





#40

Benzene

Concen: 1.772 ug/l

RT: 8.323 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VW019278.D

Acq: 22 Jun 2021 19:22

Tgt Ion: 78 Resp: 3116

Ion Ratio Lower Upper





78 100

77 26.9 19.6 29.4

Instrument :

MSVOA_W

ClientSampled :



#50

Toluene-d8

Concen: 43.427 ug/l

RT: 10.329 min Scan# 1382

Delta R.T. 0.006 min

Lab File: VW019278.D




Acq: 22 Jun 2021 19:22


Tgt Ion: 98 Resp: 67156

Ion Ratio Lower Upper

98 100

100 63.8 52.0 78.0









#52
Toluene
Concen: 19.837 ug/l
RT: 10.390 min Scan# 1392
Delta R.T. 0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 92 Resp: 21706
Ion Ratio Lower Upper
92 100
91 167.8 137.4 206.0



#62
4-Bromofluorobenzene
Concen: 39.555 ug/l
RT: 12.621 min Scan# 1758
Delta R.T. 0.006 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 95 Resp: 21393
Ion Ratio Lower Upper
95 100
174 74.0 0.0 156.2
176 73.2 0.0 160.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.634 min Scan# 1596
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion:117 Resp: 49018
Ion Ratio Lower Upper
117 100
82 61.2 44.4 66.6
119 32.9 24.6 36.8



#67
Ethyl Benzene
Concen: 20.387 ug/l
RT: 11.731 min Scan# 1612
Delta R.T. 0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 91 Resp: 38352
Ion Ratio Lower Upper
91 100
106 27.8 24.0 36.0









#68
m/p-Xylenes
Concen: 80.410 ug/l
RT: 11.835 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:106 Resp: 56473
Ion Ratio Lower Upper
106 100
91 217.8 164.2 246.2



#69
o-Xylene
Concen: 39.762 ug/l
RT: 12.170 min Scan# 1684
Delta R.T. 0.006 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion:106 Resp: 26174
Ion Ratio Lower Upper
106 100
91 225.9 109.1 327.2





#70

Styrene

Concen: 5.066 ug/l

RT: 12.176 min Scan# 1685

Delta R.T. -0.006 min

Lab File: VW019278.D

Acq: 22 Jun 2021 19:22

Tgt Ion:104 Resp: 5623

Ion Ratio Lower Upper

104 100





78 115.0 41.3 61.9#

103 109.7 44.1 66.1#

Instrument :

MSVOA_W

ClientSampled :



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.560 min Scan# 1912

Delta R.T. 0.000 min

Lab File: VW019278.D

Acq: 22 Jun 2021 19:22




Tgt Ion:152 Resp: 17519


Ion Ratio Lower Upper

152 100

115 73.7 30.3 90.8

150 154.7 0.0 356.0









#73
Isopropylbenzene
Concen: 7.231 ug/l
RT: 12.463 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:105 Resp: 9422
Ion Ratio Lower Upper
105 100
120 23.8 13.2 39.5



#74
N-ethyl acetate
Concen: 1.758 ug/l
RT: 12.250 min Scan# 1697
Delta R.T. -0.024 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 43 Resp: 485
Ion Ratio Lower Upper
43 100
70 35.1 38.4 57.6#
55 47.8 21.6 32.4#
61 0.0 21.0 31.4#









#75
1,1,2,2-Tetrachloroethane
Concen: 1.401 ug/l
RT: 12.701 min Scan# 1771
Delta R.T. -0.012 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 83 Resp: 332
Ion Ratio Lower Upper
83 100
131 40.4 5.1 15.2#
85 0.0 31.6 94.8#



#76
1,2,3-Trichloropropane
Concen: 4.446 ug/l
RT: 12.859 min Scan# 1797
Delta R.T. 0.091 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 75 Resp: 763
Ion Ratio Lower Upper
75 100
77 1449.9 0.0 0.0#









#78
n-propylbenzene
Concen: 6.458 ug/l
RT: 12.804 min Scan# 1788
Delta R.T. 0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 91 Resp: 10609
Ion Ratio Lower Upper
91 100
120 24.1 11.1 33.3



#79
2-Chlorotoluene
Concen: 4.691 ug/l
RT: 13.109 min Scan# 1838
Delta R.T. 0.219 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 91 Resp: 4405
Ion Ratio Lower Upper
91 100
126 37.7 16.5 49.5









#80
1,3,5-Trimethylbenzene
Concen: 68.440 ug/l
RT: 12.945 min Scan# 1811
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:105 Resp: 75607
Ion Ratio Lower Upper
105 100
120 46.6 24.1 72.2



#81
trans-1,4-Dichloro-2-butene
Concen: 190.115 ug/l
RT: 12.615 min Scan# 1757
Delta R.T. 0.103 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 75 Resp: 13450
Ion Ratio Lower Upper
75 100
53 1.4 81.5 122.3#
89 5.5 36.3 54.5#









#82
4-Chlorotoluene
Concen: 4.488 ug/l
RT: 13.109 min Scan# 1838
Delta R.T. 0.122 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion: 91 Resp: 4405
Ion Ratio Lower Upper
91 100
126 37.7 16.1 48.2



#83
tert-Butylbenzene
Concen: 21.087 ug/l
RT: 13.249 min Scan# 1861
Delta R.T. 0.042 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 119 Resp: 19668
Ion Ratio Lower Upper
119 100
91 97.1 34.0 102.0
134 0.0 13.0 38.9#









#84
1,2,4-Trimethylbenzene
Concen: 158.609 ug/l
RT: 13.249 min Scan# 1861
Delta R.T. 0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Instrument :
MSVOA_W
ClientSampled :


Tgt Ion:105 Resp: 173961
Ion Ratio Lower Upper
105 100
120 44.1 22.4 67.0



#85
sec-Butylbenzene
Concen: 119.911 ug/l
RT: 13.249 min Scan# 1861
Delta R.T. -0.134 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22


Tgt Ion:105 Resp: 173961
Ion Ratio Lower Upper
105 100
134 0.0 10.1 30.1#








#86
p-Isopropyltoluene
Concen: 7.119 ug/l
RT: 13.499 min Scan# 1902
Delta R.T. 0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22


Instrument :
MSVOA_W
ClientSampled :





Tgt Ion:	119	Resp:	8404
Ion	Ratio	Lower	Upper
119	100		
134	25.0	12.8	38.3
91	33.5	12.1	36.3




#88
1,4-Dichlorobenzene
Concen: 1.325 ug/l
RT: 13.579 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22




Tgt Ion:	146	Resp:	777
Ion	Ratio	Lower	Upper
146	100		
111	127.0	20.1	60.3#
148	66.3	32.4	97.2








#89
n-Butylbenzene
Concen: 52.215 ug/l
RT: 13.755 min Scan# 1944
Delta R.T. -0.067 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Instrument :
MSVOA_W
ClientSampleId :







Tgt Ion: 91 Resp: 61999
Ion Ratio Lower Upper
91 100
92 11.2 26.6 79.8#
134 5.0 12.7 38.0#



#90
Hexachloroethane
Concen: 25.639 ug/l
RT: 14.091 min Scan# 1999
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion: 117 Resp: 5486
Ion Ratio Lower Upper
117 100
201 0.0 38.4 115.2#









#92
1,2-Dibromo-3-Chloropropane
Concen: 61.095 ug/l
RT: 14.481 min Scan# 2063
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Instrument :
MSVOA_W
ClientSampleId :

Tgt Ion: 75 Resp: 2311
Ion Ratio Lower Upper
75 100
155 0.0 41.3 124.1#
157 16.1 51.2 153.6#



#95
Naphthalene
Concen: 6909.965 ug/l
RT: 15.365 min Scan# 2208
Delta R.T. -0.000 min
Lab File: VW019278.D
Acq: 22 Jun 2021 19:22

Tgt Ion:128 Resp: 4163280
Ion Ratio Lower Upper
128 100
127 13.4 10.4 15.6
129 11.3 8.5 12.7

