

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062221\
 Data File : VW019274.D
 Acq On : 22 Jun 2021 17:22
 Operator : SY/VA
 Sample : M2801-02RE
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-1ARE

Quant Time: Jun 23 05:43:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 21 12:12:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.957	168	79848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.847	114	141119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.633	117	116314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	40907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	59649	61.582	ug/l	0.00
Spiked Amount	50.000		Recovery	=	123.160%	
35) Dibromofluoromethane	7.884	113	47416	52.752	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.500%	
50) Toluene-d8	10.329	98	156674	43.542	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.080%	
62) 4-Bromofluorobenzene	12.615	95	47346	37.622	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.240%	
Target Compounds						
16) Acetone	4.129	43	5498	35.366	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062221\
Data File : VW019274.D
Acq On : 22 Jun 2021 17:22
Operator : SY/VA
Sample : M2801-02RE
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-1ARE

Quant Time: Jun 23 05:43:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 21 12:12:47 2021
Response via : Initial Calibration

