



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062221\
 Data File : VW019256.D
 Acq On : 22 Jun 2021 08:50
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W062121S.M
 Title : SW846 8260
 Last Update : Mon Jun 21 12:12:47 2021



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	45728	PASS
75	95	30	60	52.6	124464	PASS
95	95	100	100	100.0	236608	PASS
96	95	5	9	6.4	15207	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.4	178347	PASS
175	174	5	9	7.2	12893	PASS
176	174	95	101	95.6	170560	PASS
177	176	5	9	6.7	11484	PASS