

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062223\
 Data File : VW026297.D
 Acq On : 22 Jun 2023 13:10
 Operator : SY/MD
 Sample : 03325-04
 Misc : 5.08g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jun 23 01:04:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 23 01:02:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	683259	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	632595	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	277346	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	161656	16.818	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.280%
7) Chloroethane-d5	2.899	69	158046	21.965	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.840%
11) 1,1-Dichloroethene-d2	4.015	65	86010	18.147	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.600%
21) 2-Butanone-d5	7.075	46	133689	47.773	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.540%
24) Chloroform-d	7.648	84	422641	21.123	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.480%
26) 1,2-Dichloroethane-d4	8.307	65	233048	21.301	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.200%
32) Benzene-d6	8.276	84	802585	18.855	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.400%
36) 1,2-Dichloropropane-d6	9.270	67	274199	20.333	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.320%
41) Toluene-d8	10.325	98	660220	17.490	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	69.960%
43) trans-1,3-Dichloroprop...	10.575	79	91233	17.072	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.280%
47) 2-Hexanone-d5	10.922	63	86636	40.744	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.480%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	217433	22.157	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.640%
66) 1,2-Dichlorobenzene-d4	13.849	152	228575	21.532	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.120%
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	34833	3.194	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062223\
Data File : VW026297.D
Acq On : 22 Jun 2023 13:10
Operator : SY/MD
Sample : 03325-04
Misc : 5.08g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jun 23 01:04:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 23 01:02:45 2023
Response via : Initial Calibration

