

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062223\
 Data File : VW026306.D
 Acq On : 22 Jun 2023 17:53
 Operator : SY/MD
 Sample : 03382-14
 Misc : 4.33g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YBW50

Quant Time: Jun 23 01:06:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 23 01:02:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	742199	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	630403	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	250945	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	159997	15.324	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.280%
7) Chloroethane-d5	2.893	69	160238	20.501	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.000%
11) 1,1-Dichloroethene-d2	4.021	65	86934	16.886	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.560%
21) 2-Butanone-d5	7.075	46	202348	66.566	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	133.140%
24) Chloroform-d	7.648	84	451554	20.776	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.120%
26) 1,2-Dichloroethane-d4	8.307	65	268519	22.594	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.360%
32) Benzene-d6	8.276	84	860378	20.283	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.120%
36) 1,2-Dichloropropane-d6	9.270	67	296923	22.095	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.360%
41) Toluene-d8	10.325	98	705852	18.763	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.040%
43) trans-1,3-Dichloroprop...	10.575	79	99175	18.623	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.480%
47) 2-Hexanone-d5	10.922	63	125383	59.171	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.340%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	272602	27.876	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.520%
66) 1,2-Dichlorobenzene-d4	13.849	152	213381	22.215	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW062223\
Data File : VW026306.D
Acq On : 22 Jun 2023 17:53
Operator : SY/MD
Sample : 03382-14
Misc : 4.33g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YBW50

Quant Time: Jun 23 01:06:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 23 01:02:45 2023
Response via : Initial Calibration

