

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062318\
 Data File : VW003461.D
 Acq On : 23 Jun 2018 15:36
 Operator : SY/AP
 Sample : J3664-04
 Misc : 6.46G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WP021SB482-17-19-180619

Quant Time: Jun 25 03:53:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	201249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	333699	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	299665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	139394	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	78784	36.14	ug/l	0.00
Spiked Amount			Recovery	=	72.28%	
35) Dibromofluoromethane	7.88	113	65334	31.51	ug/l	0.00
Spiked Amount			Recovery	=	63.02%	
50) Toluene-d8	10.33	98	236030	28.71	ug/l	0.00
Spiked Amount			Recovery	=	57.42%	
62) 4-Bromofluorobenzene	12.63	95	82205	28.25	ug/l	0.00
Spiked Amount			Recovery	=	56.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.15	43	52637	115.13	ug/l	95
17) Carbon Disulfide	4.37	76	13995	2.60	ug/l #	93
20) Methylene Chloride	4.90	84	11206	2.27	ug/l	91
25) 2-Butanone	7.18	43	13291	23.38	ug/l	97
40) Benzene	8.33	78	89255	10.05	ug/l	99
52) Toluene	10.40	92	7873	1.38	ug/l	98
65) Chlorobenzene	11.66	112	17601	2.82	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062318\
Data File : VW003461.D
Acq On : 23 Jun 2018 15:36
Operator : SY/AP
Sample : J3664-04
Misc : 6.46G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WP021SB482-17-19-180619

Quant Time: Jun 25 03:53:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

