

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062318\
 Data File : VW003466.D
 Acq On : 23 Jun 2018 17:45
 Operator : SY/AP
 Sample : J3664-09
 Misc : 6.31G/5ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WP021SB476-27-29-180619

Quant Time: Jun 25 04:00:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	153898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	242051	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	221872	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	109294	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	73176	43.90	ug/l	0.00
Spiked Amount			Recovery	=	87.80%	
35) Dibromofluoromethane	7.89	113	60207	40.03	ug/l	0.00
Spiked Amount			Recovery	=	80.06%	
50) Toluene-d8	10.33	98	199791	33.50	ug/l	0.00
Spiked Amount			Recovery	=	67.00%	
62) 4-Bromofluorobenzene	12.63	95	69273	32.82	ug/l	0.00
Spiked Amount			Recovery	=	65.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.36	62	27566	16.09	ug/l	99
16) Acetone	4.16	43	11270	32.24	ug/l #	89
20) Methylene Chloride	4.91	84	8151	1.96	ug/l	91
21) trans-1,2-Dichloroethene	5.42	96	8029	5.30	ug/l	95
27) cis-1,2-Dichloroethene	7.17	96	171889	103.20	ug/l	92
44) Trichloroethene	9.09	130	10055	5.89	ug/l	91
65) Chlorobenzene	11.66	112	6481	1.40	ug/l	96
91) 1,2-Dichlorobenzene	13.88	146	17249	5.41	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062318\
Data File : VW003466.D
Acq On : 23 Jun 2018 17:45
Operator : SY/AP
Sample : J3664-09
Misc : 6.31G/5ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WP021SB476-27-29-180619

Quant Time: Jun 25 04:00:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

