

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW062420\
 Data File : VW015676.D
 Acq On : 24 Jun 2020 12:16
 Operator : SY/VA
 Sample : L3020-16
 Misc : 6.93G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 MW-27(13-15)

Quant Time: Jul 01 12:20:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 23 13:20:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	165418	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	295429	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	270967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	119801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	104811	56.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.90%	
35) Dibromofluoromethane	7.88	113	94531	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	10.32	98	370866	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
62) 4-Bromofluorobenzene	12.62	95	124412	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
Target Compounds						
					Qvalue	
16) Acetone	4.15	43	4941	14.122	ug/l	91
39) Methylcyclohexane	9.34	83	4855	1.304	ug/l #	84
67) Ethyl Benzene	11.73	91	28449	2.676	ug/l	98
78) n-propylbenzene	12.80	91	15088	1.372	ug/l	100
80) 1,3,5-Trimethylbenzene	12.94	105	23337	3.014	ug/l	99
84) 1,2,4-Trimethylbenzene	13.26	105	44370	5.746	ug/l	99
95) Naphthalene	15.37	128	29197	7.643	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW062420\
Data File : VW015676.D
Acq On : 24 Jun 2020 12:16
Operator : SY/VA
Sample : L3020-16
Misc : 6.93G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
MW-27(13-15)

Quant Time: Jul 01 12:20:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 23 13:20:19 2020
Response via : Initial Calibration

