

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062420\
 Data File : VW015713.D
 Acq On : 25 Jun 2020 05:27
 Operator : SY/VA
 Sample : L3111-06
 Misc : 8.06G/5ML/MSVOA W/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-6-D2

Quant Time: Jun 25 06:30:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 23 13:20:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	88326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	149744	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	140458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	52782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	64672	65.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.62%	
35) Dibromofluoromethane	7.88	113	48602	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
50) Toluene-d8	10.32	98	185610	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	12.62	95	59484	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
16) Acetone	4.12	43	2169	11.610	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062420\
Data File : VW015713.D
Acq On : 25 Jun 2020 05:27
Operator : SY/VA
Sample : L3111-06
Misc : 8.06G/5ML/MSVOA W/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-6-D2

Quant Time: Jun 25 06:30:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 23 13:20:19 2020
Response via : Initial Calibration

