

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW062518\
 Data File : VW003546.D
 Acq On : 25 Jun 2018 14:11
 Operator : SY/AP
 Sample : J3621-14
 Misc : 8.07G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DAZ72

Quant Time: Jun 26 03:05:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 26 03:05:07 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	244098	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	232173	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	98975	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	68702	20.75	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.00%
7) Chloroethane-d5	2.89	69	53927	19.52	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.08%
10) 1,1-Dichloroethene-d2	4.01	63	106500	16.20	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.80%
20) 2-Butanone-d5	7.09	46	52680	50.50	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.00%
24) Chloroform-d	7.65	84	148266	23.09	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.36%
26) 1,2-Dichloroethane-d4	8.31	65	93341	24.46	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.84%
29) Benzene-d6	8.28	84	271950	22.08	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.32%
33) 1,2-Dichloropropane-d6	9.28	67	91385	23.44	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.76%
37) Toluene-d8	10.33	98	237081	20.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.52%
38) trans-1,3-Dichloropropene-	10.58	79	35753	19.53	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.12%
39) 2-Hexanone-d5	10.93	63	35049	43.97	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.94%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	85368	23.53	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.12%
61) 1,2-Dichlorobenzene-d4	13.86	152	91619	24.51	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.04%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW062518\
Data File : VW003546.D
Acq On : 25 Jun 2018 14:11
Operator : SY/AP
Sample : J3621-14
Misc : 8.07G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DAZ72

Quant Time: Jun 26 03:05:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 26 03:05:07 2018
Response via : Initial Calibration

