

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062619\
 Data File : VW010998.D
 Acq On : 26 Jun 2019 12:24
 Operator : SY/VA
 Sample : K3461-04MS
 Misc : 4.73G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YANZ3MS

Manual Integrations
 APPROVED

MMDadoda
 6/29/2019 3:53:24 AM

Quant Time: Jun 27 08:06:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062419S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 27 07:30:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	50912	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	50022	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	21323	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.34	65	12898	19.54	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.16%
7) Chloroethane-d5	2.88	69	11089	21.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.20%
10) 1,1-Dichloroethene-d2	4.01	63	31390	20.13	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	80.52%
20) 2-Butanone-d5	7.08	46	18666	60.23	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	120.46%
24) Chloroform-d	7.65	84	32241	23.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.92%
26) 1,2-Dichloroethane-d4	8.31	65	23608	27.74	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	110.96%
29) Benzene-d6	8.27	84	63396	21.11	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.44%
33) 1,2-Dichloropropane-d6	9.27	67	23069	23.25	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.00%
37) Toluene-d8	10.32	98	52449	19.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.44%
38) trans-1,3-Dichloropropene-	10.58	79	8404	19.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.56%
39) 2-Hexanone-d5	10.92	63	11862	53.19	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.38%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	23039	28.17	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.68%
61) 1,2-Dichlorobenzene-d4	13.85	152	19795	24.59	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.36%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	4.03	96	12893	18.092	ug/L	94
13) Acetone	4.13	43	2312m	7.790	ug/L	
16) Methylene chloride	4.90	84	2154m	2.857	ug/L	
34) Benzene	8.32	78	65580	19.951	ug/L	100
35) Trichloroethene	9.09	95	15774	18.923	ug/L	99
44) Toluene	10.39	91	65014	18.915	ug/L	98
52) Chlorobenzene	11.66	112	41121	19.881	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062619\
Data File : VW010998.D
Acq On : 26 Jun 2019 12:24
Operator : SY/VA
Sample : K3461-04MS
Misc : 4.73G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YANZ3MS

Manual Integrations
APPROVED

MMDadoda
6/29/2019 3:53:24 AM

Quant Time: Jun 27 08:06:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062419S.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 27 07:30:59 2019
Response via : Initial Calibration

