

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062620\
 Data File : VW015741.D
 Acq On : 26 Jun 2020 11:31
 Operator : SY/VA
 Sample : VW0626SBS01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 27 01:48:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 23 13:20:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	195971	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	309868	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	283244	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	149207	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	102331	46.93	ug/l	0.00
Spiked Amount			Recovery	=	93.86%	
35) Dibromofluoromethane	7.89	113	91343	46.86	ug/l	0.00
Spiked Amount			Recovery	=	93.72%	
50) Toluene-d8	10.32	98	367913	47.96	ug/l	0.00
Spiked Amount			Recovery	=	95.92%	
62) 4-Bromofluorobenzene	12.62	95	134356	47.56	ug/l	0.00
Spiked Amount			Recovery	=	95.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	2.01	85	22677	19.407	ug/l	99
3) Chloromethane	2.21	50	27595	20.111	ug/l	98
4) Vinyl Chloride	2.37	62	45049	20.308	ug/l	100
5) Bromomethane	2.78	94	34698	21.282	ug/l	95
6) Chloroethane	2.93	64	30485	20.883	ug/l	99
7) Trichlorofluoromethane	3.26	101	33684	20.224	ug/l	100
8) Diethyl Ether	3.68	74	19453	20.223	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	4.06	101	38859	19.793	ug/l	99
10) Methyl Iodide	4.28	142	54582	19.662	ug/l	99
11) Tert butyl alcohol	5.20	59	11700	73.288	ug/l	99
12) 1,1-Dichloroethene	4.04	96	37672	19.707	ug/l	98
13) Acrolein	3.90	56	13136	86.724	ug/l	99
14) Allyl chloride	4.67	41	58872	18.982	ug/l	99
15) Acrylonitrile	5.38	53	38060	95.901	ug/l	99
16) Acetone	4.13	43	39388	95.023	ug/l	96
17) Carbon Disulfide	4.38	76	98892	17.974	ug/l	99
18) Methyl Acetate	4.68	43	19730	20.249	ug/l	100
19) Methyl tert-butyl Ether	5.43	73	57639	20.035	ug/l	98
20) Methylene Chloride	4.92	84	51199	22.881	ug/l	96
21) trans-1,2-Dichloroethene	5.43	96	42819	19.537	ug/l	96
22) Diisopropyl ether	6.32	45	121271	19.660	ug/l	96
23) Vinyl Acetate	6.26	43	343972	95.624	ug/l	99
24) 1,1-Dichloroethane	6.22	63	76080	19.594	ug/l	100
25) 2-Butanone	7.18	43	52734	94.130	ug/l	98
26) 2,2-Dichloropropane	7.17	77	49368	18.957	ug/l	99
27) cis-1,2-Dichloroethene	7.17	96	45861	19.451	ug/l	98
28) Bromochloromethane	7.52	49	29987	20.077	ug/l	99
29) Tetrahydrofuran	7.54	42	31274	94.997	ug/l	99
30) Chloroform	7.68	83	77837	19.547	ug/l	100
31) Cyclohexane	7.96	56	70301	18.717	ug/l	95
32) 1,1,1-Trichloroethane	7.88	97	65371	19.258	ug/l	99
36) 1,1-Dichloropropene	8.09	75	63403	19.672	ug/l	100
37) Ethyl Acetate	7.26	43	23051	18.732	ug/l	99
38) Carbon Tetrachloride	8.07	117	59609	19.080	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062620\
 Data File : VW015741.D
 Acq On : 26 Jun 2020 11:31
 Operator : SY/VA
 Sample : VW0626SBS01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 27 01:48:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 23 13:20:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.34	83	74862	19.177	ug/l	98
40) Benzene	8.32	78	171671	19.590	ug/l	100
41) Methacrylonitrile	7.49	41	12990	18.393	ug/l	98
42) 1,2-Dichloroethane	8.40	62	53624	19.780	ug/l	100
43) Isopropyl Acetate	8.43	43	45980	18.724	ug/l	98
44) Trichloroethene	9.10	130	45804	19.391	ug/l	95
45) 1,2-Dichloropropane	9.37	63	41604	19.423	ug/l	99
46) Dibromomethane	9.46	93	21743	19.619	ug/l	98
47) Bromodichloromethane	9.65	83	54755	18.940	ug/l	99
48) Methyl methacrylate	9.44	41	20654	18.396	ug/l	94
49) 1,4-Dioxane	9.46	88	5369	369.286	ug/l	95
51) 4-Methyl-2-Pentanone	10.21	43	112614	95.090	ug/l	100
52) Toluene	10.39	92	113559	20.068	ug/l	98
53) t-1,3-Dichloropropene	10.61	75	50946	18.242	ug/l	99
54) cis-1,3-Dichloropropene	10.07	75	62720	18.799	ug/l	98
55) 1,1,2-Trichloroethane	10.79	97	30342	19.691	ug/l	98
56) Ethyl methacrylate	10.65	69	36261	18.912	ug/l	96
57) 1,3-Dichloropropane	10.93	76	53654	19.640	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.93	63	94289	103.921	ug/l	99
59) 2-Hexanone	10.97	43	78397	97.236	ug/l	99
60) Dibromochloromethane	11.13	129	33308	18.549	ug/l	96
61) 1,2-Dibromoethane	11.24	107	28907	19.626	ug/l	100
64) Tetrachloroethene	10.87	164	39123	19.661	ug/l	97
65) Chlorobenzene	11.66	112	119016	19.929	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.73	131	40430	19.308	ug/l	98
67) Ethyl Benzene	11.73	91	220389	19.830	ug/l	99
68) m/p-Xylenes	11.84	106	168005	40.257	ug/l	98
69) o-Xylene	12.17	106	75673	19.721	ug/l	99
70) Styrene	12.18	104	129282	19.720	ug/l	98
71) Bromoform	12.35	173	17581	18.249	ug/l #	100
73) Isopropylbenzene	12.46	105	216899	19.056	ug/l	99
74) N-amyl acetate	12.27	43	44407	18.404	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.72	83	33910	18.623	ug/l	99
76) 1,2,3-Trichloropropane	12.77	75	43571	32.014	ug/l #	1
77) Bromobenzene	12.75	156	46894	18.477	ug/l	96
78) n-propylbenzene	12.80	91	261151	19.061	ug/l	99
79) 2-Chlorotoluene	12.89	91	150443	19.502	ug/l	98
80) 1,3,5-Trimethylbenzene	12.94	105	184077	19.086	ug/l	98
81) trans-1,4-Dichloro-2-buten	12.51	75	8658	16.031	ug/l	87
82) 4-Chlorotoluene	12.99	91	156490	19.175	ug/l	99
83) tert-Butylbenzene	13.21	119	156571	19.214	ug/l	98
84) 1,2,4-Trimethylbenzene	13.25	105	187310	19.475	ug/l	100
85) sec-Butylbenzene	13.38	105	225671	19.377	ug/l	99
86) p-Isopropyltoluene	13.50	119	204869	19.361	ug/l	99
87) 1,3-Dichlorobenzene	13.50	146	96327	19.119	ug/l	98
88) 1,4-Dichlorobenzene	13.58	146	95281	19.095	ug/l	97
89) n-Butylbenzene	13.82	91	198061	19.499	ug/l	98
90) Hexachloroethane	14.10	117	33977	18.044	ug/l	99
91) 1,2-Dichlorobenzene	13.87	146	83989	19.280	ug/l	100
92) 1,2-Dibromo-3-Chloropropan	14.49	75	5138	17.621	ug/l	97


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062620\
Data File : VW015741.D
Acq On : 26 Jun 2020 11:31
Operator : SY/VA
Sample : VW0626SBSD01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jun 27 01:48:22 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 23 13:20:19 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	15.13	180	53075	18.450	ug/l	98
94) Hexachlorobutadiene	15.24	225	33743	18.412	ug/l	99
95) Naphthalene	15.37	128	86941	18.273	ug/l	100
96) 1,2,3-Trichlorobenzene	15.56	180	45509	18.374	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 992

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31


Instrument :
MSVOA_W
ClientSampled :

Tot Ion:168 Resp: 195971

Ion Ratio Lower Upper

168 100

99 59.3 48.6 73.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

100000

80000



60000

40000

20000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 19.407 ug/l

RT: 2.01 min Scan# 18

Delta R.T. 0.00 min

Lab File: VW015741.D


Acq: 26 Jun 2020 11:31

Tgt Ion: 85 Resp: 22677

Ion Ratio Lower Upper

85 100

87 32.5 15.9 47.6



Abundance Ion 84.90 (84.60 to 85.60): VW

Ion 86.90 (86.60 to 87.60): VW

2.01

8000


6000


4000

2000

0

Time-->









#3
Chloromethane
Concen: 20.111 ug/l
RT: 2.21 min Scan# 51
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 50 Resp: 27595
Ion Ratio Lower Upper
50 100
52 31.3 26.0 39.0



#4
Vinyl Chloride
Concen: 20.308 ug/l
RT: 2.37 min Scan# 76
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 62 Resp: 45049
Ion Ratio Lower Upper
62 100
64 31.6 25.1 37.7





#5

Bromomethane

Concen: 21.282 ug/l

RT: 2.78 min Scan# 144

Delta R.T. -0.01 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W





ClientSampleId :

Tot Ion: 94 Resp: 34698

Ion Ratio Lower Upper

94 100

96 97.2 73.8 110.6



#6

Chloroethane

Concen: 20.883 ug/l

RT: 2.93 min Scan# 168

Delta R.T. -0.01 min

Lab File: VW015741.D




Acq: 26 Jun 2020 11:31


Tgt Ion: 64 Resp: 30485

Ion Ratio Lower Upper

64 100

66 30.8 25.3 37.9






#7




Trichlorofluoromethane
Concen: 20.224 ug/l
RT: 3.26 min Scan# 223
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Instrument :
MSVOA_W
ClientSampled :

Tot Ion:101 Resp: 33684
Ion Ratio Lower Upper
101 100
103 65.9 52.6 78.8




Abundance Ion 100.90 (100.60 to 101.60): V
Ion 102.80 (102.50 to 103.50): V





#8


Diethyl Ether
Concen: 20.223 ug/l
RT: 3.68 min Scan# 291
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 74 Resp: 19453
Ion Ratio Lower Upper
74 100
45 97.7 48.2 144.6



Abundance Ion 74.00 (73.70 to 74.70): VW
Ion 45.00 (44.70 to 45.70): VW





#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.793 ug/l

RT: 4.06 min Scan# 354

Delta R.T. -0.01 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :
MSVOA_W
ClientSampleId :


Tot Ion:101 Resp: 38859

Ion Ratio Lower Upper




101 100

85 45.0 36.6 55.0

151 75.2 60.4 90.6



Abundance Ion 100.90 (100.60 to 101.60): V
Ion 84.90 (84.60 to 85.60): VW
Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 19.662 ug/l

RT: 4.28 min Scan# 389

Delta R.T. 0.01 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31


Tgt Ion:142 Resp: 54582

Ion Ratio Lower Upper



142 100


127 45.0 36.4 54.6

141 14.4 11.6 17.4



Abundance Ion 141.80 (141.50 to 142.50): V
Ion 126.80 (126.50 to 127.50): V
Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 73.288 ug/l
 RT: 5.20 min Scan# 541
 Delta R.T. 0.02 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Instrument :
 MSVOA_W
 ClientSampled :




Tot Ion: 59 Resp: 11700
 Ion Ratio Lower Upper
 59 100
 57 10.1 8.4 12.6




#12

1,1-Dichloroethene
 Concen: 19.707 ug/l
 RT: 4.04 min Scan# 350
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 96 Resp: 37672
 Ion Ratio Lower Upper
 96 100
 61 172.5 136.7 205.1
 98 64.4 49.2 73.8





#13

Acrolein

Concen: 86.724 ug/l

RT: 3.90 min Scan# 327

Delta R.T. 0.01 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31





Instrument :
MSVOA_W
ClientSampled :

Tot Ion: 56 Resp: 13136

Ion Ratio Lower Upper

56 100

55 69.7 56.2 84.2



#14

Allyl chloride

Concen: 18.982 ug/l

RT: 4.67 min Scan# 454

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31




Tgt Ion: 41 Resp: 58872


Ion Ratio Lower Upper

41 100

39 81.5 64.6 97.0

76 35.4 29.1 43.7





#15

Acrylonitrile

Concen: 95.901 ug/l

RT: 5.38 min Scan# 570

Delta R.T. 0.01 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :
MSVOA_W
ClientSampled :


Tot Ion: 53 Resp: 38060

Ion Ratio Lower Upper




53 100

52 83.1 65.8 98.6

51 39.1 30.7 46.1



Abundance Ion 53.00 (52.70 to 53.70): VW
Ion 52.00 (51.70 to 52.70): VW
Ion 51.00 (50.70 to 51.70): VW



#16

Acetone

Concen: 95.023 ug/l

RT: 4.13 min Scan# 366

Delta R.T. 0.01 min

Lab File: VW015741.D


Acq: 26 Jun 2020 11:31

Tgt Ion: 43 Resp: 39388



Ion Ratio Lower Upper


43 100

58 26.2 22.6 33.8



Abundance Ion 43.00 (42.70 to 43.70): VW
Ion 58.00 (57.70 to 58.70): VW





#17

Carbon Disulfide

Concen: 17.974 ug/l

RT: 4.38 min Scan# 407

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W


ClientSampleId :

Tot Ion: 76 Resp: 98892

Ion Ratio Lower Upper




76 100

78 8.8 7.2 10.8



Abundance Ion 75.90 (75.60 to 76.60): VW

Ion 77.90 (77.60 to 78.60): VW



#18

Methyl Acetate

Concen: 20.249 ug/l

RT: 4.68 min Scan# 455

Delta R.T. 0.01 min

Lab File: VW015741.D


Acq: 26 Jun 2020 11:31

Tgt Ion: 43 Resp: 19730

Ion Ratio Lower Upper



43 100


74 23.7 19.0 28.6



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 74.00 (73.70 to 74.70): VW









#19
Methvl tert-butvl Ether
Concen: 20.035 ug/l
RT: 5.43 min Scan# 579
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 73 Resp: 57639
Ion Ratio Lower Upper
73 100
57 23.4 18.1 27.1



#20
Methylene Chloride
Concen: 22.881 ug/l
RT: 4.92 min Scan# 495
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 84 Resp: 51199
Ion Ratio Lower Upper
84 100
49 118.0 99.3 148.9
51 38.7 31.8 47.6
86 61.1 51.7 77.5





#21

trans-1,2-Dichloroethene

Concen: 19.537 ug/l

RT: 5.43 min Scan# 578

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampled :


Tot Ion: 96 Resp: 42819

Ion Ratio Lower Upper




96 100

61 139.6 116.2 174.2

98 63.1 51.4 77.0



Abundance Ion 95.90 (95.60 to 96.60): VW
Ion 60.90 (60.60 to 61.60): VW
Ion 97.90 (97.60 to 98.60): VW



#22

Diisopropyl ether

Concen: 19.660 ug/l

RT: 6.32 min Scan# 724

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Tgt Ion: 45 Resp: 121271


Ion Ratio Lower Upper

45 100



43 58.2 43.9 65.9


87 28.5 21.2 31.8

59 12.3 9.2 13.8



Abundance Ion 45.00 (44.70 to 45.70): VW
Ion 43.00 (42.70 to 43.70): VW
Ion 87.00 (86.70 to 87.70): VW
Ion 59.00 (58.70 to 59.70): VW









#23
 Vinyl Acetate
 Concen: 95.624 ug/l
 RT: 6.26 min Scan# 715
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampled :


Tot Ion: 43 Resp: 343972
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.3 12.5



#24
 1,1-Dichloroethane
 Concen: 19.594 ug/l
 RT: 6.22 min Scan# 708
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 63 Resp: 76080
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.5 10.5
 100 4.1 2.1 6.3









#25
2-Butanone
Concen: 94.130 ug/l
RT: 7.18 min Scan# 866
Delta R.T. 0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampled :


Tot Ion: 43 Resp: 52734
Ion Ratio Lower Upper
43 100
72 25.4 21.1 31.7



#26
2,2-Dichloropropane
Concen: 18.957 ug/l
RT: 7.17 min Scan# 864
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 77 Resp: 49368
Ion Ratio Lower Upper
77 100
97 21.7 11.2 33.5









#27
 cis-1,2-Dichloroethene
 Concen: 19.451 ug/l
 RT: 7.17 min Scan# 864
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Instrument :
 MSVOA_W
 ClientSampled :

Tot Ion: 96 Resp: 45861
 Ion Ratio Lower Upper
 96 100
 61 150.6 0.0 296.8
 98 65.8 0.0 128.0




Abundance Ion 95.90 (95.60 to 96.60): VW
 Ion 60.90 (60.60 to 61.60): VW
 Ion 97.90 (97.60 to 98.60): VW





#28
 Bromochloromethane
 Concen: 20.077 ug/l
 RT: 7.52 min Scan# 921
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 49 Resp: 29987
 Ion Ratio Lower Upper
 49 100
 129 1.2 0.0 2.2
 130 75.1 61.0 91.4



Abundance Ion 48.90 (48.60 to 49.60): VW
 Ion 128.80 (128.50 to 129.50): V
 Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 94.997 ug/l
RT: 7.54 min Scan# 924
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampled :


Tot Ion: 42 Resp: 31274
Ion Ratio Lower Upper
42 100
72 42.9 35.0 52.6
71 40.7 32.6 49.0



#30
Chloroform
Concen: 19.547 ug/l
RT: 7.68 min Scan# 948
Delta R.T. 0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 83 Resp: 77837
Ion Ratio Lower Upper
83 100
85 64.0 51.0 76.4





#31

Cyclohexane

Concen: 18.717 ug/l

RT: 7.96 min Scan# 993

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampled :


Tot Ion: 56 Resp: 70301

Ion Ratio Lower Upper




56 100

69 38.9 26.2 39.4

84 91.2 70.4 105.6



Abundance Ion 56.00 (55.70 to 56.70): VW
Ion 69.00 (68.70 to 69.70): VW
Ion 84.00 (83.70 to 84.70): VW



#32

1,1,1-Trichloroethane

Concen: 19.258 ug/l

RT: 7.88 min Scan# 980

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31


Tgt Ion: 97 Resp: 65371

Ion Ratio Lower Upper



97 100

99 63.9 51.9 77.9

61 47.7 37.3 55.9



Abundance Ion 96.90 (96.60 to 97.60): VW
Ion 98.90 (98.60 to 99.60): VW
Ion 60.90 (60.60 to 61.60): VW









#33
 1,2-Dichloroethane-d4
 Concen: 46.931 ug/l
 RT: 8.31 min Scan# 1051
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampleId :


Tot Ion: 65 Resp: 102331
 Ion Ratio Lower Upper
 65 100
 67 50.6 0.0 102.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.85 min Scan# 1139
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 114 Resp: 309868
 Ion Ratio Lower Upper
 114 100
 63 21.3 0.0 42.4
 88 16.2 0.0 32.6









#35
 Dibromofluoromethane
 Concen: 46.863 ug/l
 RT: 7.89 min Scan# 982
 Delta R.T. 0.01 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampled :


Tot Ion:113 Resp: 91343
 Ion Ratio Lower Upper
 113 100
 111 104.3 82.3 123.5
 192 19.9 16.4 24.6



#36
 1,1-Dichloropropene
 Concen: 19.672 ug/l
 RT: 8.09 min Scan# 1014
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 75 Resp: 63403
 Ion Ratio Lower Upper
 75 100
 110 34.0 17.0 51.0
 77 31.1 24.6 36.8









#37
Ethyl Acetate
Concen: 18.732 ug/l
RT: 7.26 min Scan# 879
Delta R.T. 0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 43 Resp: 23051
Ion Ratio Lower Upper
43 100
61 13.5 10.4 15.6
70 10.5 8.4 12.6



#38
Carbon Tetrachloride
Concen: 19.080 ug/l
RT: 8.07 min Scan# 1012
Delta R.T. 0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 117 Resp: 59609
Ion Ratio Lower Upper
117 100
119 93.3 73.2 109.8
121 30.4 23.9 35.9









#39
Methvlcvclohexane
Concen: 19.177 ug/l
RT: 9.34 min Scan# 1219
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampled :


Tot Ion: 83 Resp: 74862
Ion Ratio Lower Upper
83 100
55 77.8 62.8 94.2
98 47.4 39.9 59.9



#40
Benzene
Concen: 19.590 ug/l
RT: 8.32 min Scan# 1053
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 78 Resp: 171671
Ion Ratio Lower Upper
78 100
77 24.3 19.3 28.9









#41
Methacrylonitrile
Concen: 18.393 ug/l
RT: 7.49 min Scan# 916
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Instrument :
MSVOA_W
ClientSampleId :

Tot Ion:	41	Resp:	12990
Ion	Ratio	Lower	Upper
41	100		
39	62.3	52.0	78.0
67	71.1	57.5	86.3
52	33.4	26.5	39.7




Abundance Ion 41.00 (40.70 to 41.70): VW
Ion 39.00 (38.70 to 39.70): VW
Ion 67.00 (66.70 to 67.70): VW
Ion 52.00 (51.70 to 52.70): VW






#42
1,2-Dichloroethane
Concen: 19.780 ug/l
RT: 8.40 min Scan# 1066
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:	62	Resp:	53624
Ion	Ratio	Lower	Upper
62	100		
98	8.8	0.0	17.4



Abundance Ion 61.90 (61.60 to 62.60): VW
Ion 97.90 (97.60 to 98.60): VW





#43

Isopropyl Acetate

Concen: 18.724 ug/l

RT: 8.43 min Scan# 1070

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampled :





Tot Ion: 43 Resp: 45980

Ion Ratio Lower Upper

43 100

61 32.0 24.7 37.1

87 13.0 10.2 15.2



#44

Trichloroethene

Concen: 19.391 ug/l

RT: 9.10 min Scan# 1180

Delta R.T. 0.00 min

Lab File: VW015741.D




Acq: 26 Jun 2020 11:31


Tgt Ion: 130 Resp: 45804

Ion Ratio Lower Upper

130 100

95 102.3 0.0 194.4









#45
 1,2-Dichloropropane
 Concen: 19.423 ug/l
 RT: 9.37 min Scan# 1225
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampled :


Tot Ion: 63 Resp: 41604
 Ion Ratio Lower Upper
 63 100
 65 32.3 25.3 37.9



#46
 Dibromomethane
 Concen: 19.619 ug/l
 RT: 9.46 min Scan# 1239
 Delta R.T. -0.01 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 93 Resp: 21743
 Ion Ratio Lower Upper
 93 100
 95 82.6 65.7 98.5
 174 95.7 79.0 118.6









#47
Bromodichloromethane
Concen: 18.940 ug/l
RT: 9.65 min Scan# 1270
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion	83	Resp	54755
Ion Ratio	Lower	Upper	
83	100		
85	62.9	50.8	76.2
127	8.1	6.3	9.5



#48
Methyl methacrylate
Concen: 18.396 ug/l
RT: 9.44 min Scan# 1236
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion	41	Resp	20654
Ion Ratio	Lower	Upper	
41	100		
69	89.7	68.2	102.2
39	67.1	49.8	74.6





#49

1,4-Dioxane

Concen: 369.286 ug/l

RT: 9.46 min Scan# 1240

Delta R.T. 0.02 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampled :


Tot Ion: 88 Resp: 5369

Ion Ratio Lower Upper

88 100

43 36.3 30.1 45.1




58 75.2 56.1 84.1



Abundance Ion 88.00 (87.70 to 88.70): VW

Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW



#50

Toluene-d8

Concen: 47.963 ug/l

RT: 10.32 min Scan# 1381

Delta R.T. -0.01 min

Lab File: VW015741.D


Acq: 26 Jun 2020 11:31

Tgt Ion: 98 Resp: 367913

Ion Ratio Lower Upper



98 100

100 65.3 52.8 79.2



Abundance Ion 98.00 (97.70 to 98.70): VW

Ion 100.00 (99.70 to 100.70): VW









#51
4-Methyl-2-Pentanone
Concen: 95.090 ug/l
RT: 10.21 min Scan# 1363
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 43 Resp: 112614
Ion Ratio Lower Upper
43 100
58 39.1 31.2 46.8



#52
Toluene
Concen: 20.068 ug/l
RT: 10.39 min Scan# 1392
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 92 Resp: 113559
Ion Ratio Lower Upper
92 100
91 171.8 135.5 203.3





#53

t-1,3-Dichloropropene

Concen: 18.242 ug/l

RT: 10.61 min Scan# 1428

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W





ClientSampleId :

Tot Ion: 75 Resp: 50946

Ion Ratio Lower Upper

75 100

77 32.2 25.3 37.9



#54

cis-1,3-Dichloropropene

Concen: 18.799 ug/l

RT: 10.07 min Scan# 1340

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31




Tgt Ion: 75 Resp: 62720


Ion Ratio Lower Upper

75 100

77 32.1 25.4 38.2

39 59.1 45.8 68.6









#55
 1,1,2-Trichloroethane
 Concen: 19.691 ug/l
 RT: 10.79 min Scan# 1457
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampled :


Tot Ion:	97	Resp:	30342
Ion	Ratio	Lower	Upper
97	100		
83	81.7	66.3	99.5
85	52.2	42.5	63.7
99	63.0	48.3	72.5



#56
 Ethyl methacrylate
 Concen: 18.912 ug/l
 RT: 10.65 min Scan# 1434
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion:	69	Resp:	36261
Ion	Ratio	Lower	Upper
69	100		
41	70.4	55.8	83.6
39	42.4	29.5	44.3









#57
 1,3-Dichloropropane
 Concen: 19.640 ug/l
 RT: 10.93 min Scan# 1481
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampleId :


Tot Ion: 76 Resp: 53654
 Ion Ratio Lower Upper
 76 100
 78 33.0 26.2 39.2



#58
 2-Chloroethyl Vinyl ether
 Concen: 103.921 ug/l
 RT: 9.93 min Scan# 1316
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion: 63 Resp: 94289
 Ion Ratio Lower Upper
 63 100
 106 28.5 23.3 34.9









#59
2-Hexanone
Concen: 97.236 ug/l
RT: 10.97 min Scan# 1487
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 43 Resp: 78397
Ion Ratio Lower Upper
43 100
58 52.0 26.5 79.4



#60
Dibromochloromethane
Concen: 18.549 ug/l
RT: 11.13 min Scan# 1513
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 129 Resp: 33308
Ion Ratio Lower Upper
129 100
127 79.3 38.0 114.1





#61

1,2-Dibromoethane

Concen: 19.626 ug/l

RT: 11.24 min Scan# 1531

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W


ClientSampleId :

Tot Ion: 107 Resp: 28907

Ion Ratio Lower Upper




107 100

109 93.9 74.9 112.3



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 47.561 ug/l

RT: 12.62 min Scan# 1758

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31


Tgt Ion: 95 Resp: 134356

Ion Ratio Lower Upper

95 100

174 77.3 0.0 157.0



176 76.2 0.0 151.4




Abundance Ion 94.90 (94.60 to 95.60): VW

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Instrument :
MSVOA_W
ClientSampleId :

Tot Ion:117 Resp: 283244
Ion Ratio Lower Upper
117 100
82 56.4 45.1 67.7
119 31.1 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#64
Tetrachloroethene
Concen: 19.661 ug/l
RT: 10.87 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:164 Resp: 39123
Ion Ratio Lower Upper
164 100
166 121.7 99.9 149.9
129 89.5 74.2 111.2
131 91.0 71.4 107.0



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 19.929 ug/l
RT: 11.66 min Scan# 1600
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion:112 Resp: 119016
Ion Ratio Lower Upper
112 100
114 32.9 24.9 37.3



#66
1,1,1,2-Tetrachloroethane
Concen: 19.308 ug/l
RT: 11.73 min Scan# 1612
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:131 Resp: 40430
Ion Ratio Lower Upper
131 100
133 96.3 46.5 139.5
119 65.2 32.5 97.5





#67

Ethyl Benzene

Concen: 19.830 ug/l

RT: 11.73 min Scan# 1612

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W





ClientSampled :

Tot Ion: 91 Resp: 220389

Ion Ratio Lower Upper

91 100

106 30.9 24.1 36.1



#68

m/p-Xylenes

Concen: 40.257 ug/l

RT: 11.84 min Scan# 1630

Delta R.T. 0.00 min

Lab File: VW015741.D




Acq: 26 Jun 2020 11:31


Tgt Ion: 106 Resp: 168005

Ion Ratio Lower Upper

106 100

91 205.8 162.6 244.0









#69
o-Xylene
Concen: 19.721 ug/l
RT: 12.17 min Scan# 1684
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampled :


Tot Ion:106 Resp: 75673
Ion Ratio Lower Upper
106 100
91 218.2 108.3 324.9



#70
Styrene
Concen: 19.720 ug/l
RT: 12.18 min Scan# 1686
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:104 Resp: 129282
Ion Ratio Lower Upper
104 100
78 52.5 40.9 61.3
103 54.8 43.1 64.7





#71

Bromoform

Concen: 18.249 ug/l

RT: 12.35 min Scan# 1714

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampled :


Tot Ion:173 Resp: 17581

Ion Ratio Lower Upper




173 100

175 46.8 23.5 70.6

254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.56 min Scan# 1912

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31


Tgt Ion:152 Resp: 149207

Ion Ratio Lower Upper



152 100


115 58.6 30.1 90.5

150 161.2 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 19.056 ug/l

RT: 12.46 min Scan# 1732

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W


ClientSampled :

Tot Ion:105 Resp: 216899

Ion Ratio Lower Upper

105 100

120 26.2 12.9 38.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

12.46



150000

100000

50000

0

Time-->



#74

N-ethyl acetate

Concen: 18.404 ug/l

RT: 12.27 min Scan# 1701

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Tgt Ion: 43 Resp: 44407


Ion Ratio Lower Upper

43 100

70 42.5 33.6 50.4

55 24.1 19.7 29.5

61 23.3 18.7 28.1



Abundance Ion 43.00 (42.70 to 43.70): VV

Ion 70.00 (69.70 to 70.70): VV

Ion 55.00 (54.70 to 55.70): VV

Ion 61.00 (60.70 to 61.70): VV

40000


30000


20000

10000

0

Time-->





#75

1,1,2,2-Tetrachloroethane

Concen: 18.623 ug/l

RT: 12.72 min Scan# 1774

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampleId :


Tot Ion: 83 Resp: 33910

Ion Ratio Lower Upper

83 100

131 11.1 5.5 16.4




85 64.4 31.9 95.7



Abundance Ion 82.90 (82.60 to 83.60): VW

Ion 130.80 (130.50 to 131.50): VW

Ion 84.90 (84.60 to 85.60): VW



#76

1,2,3-Trichloropropane

Concen: 32.014 ug/l

RT: 12.77 min Scan# 1782

Delta R.T. 0.00 min

Lab File: VW015741.D


Acq: 26 Jun 2020 11:31

Tgt Ion: 75 Resp: 43571

Ion Ratio Lower Upper



75 100


77 0.0 179.1 537.1#



Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 77.00 (76.70 to 77.70): VW





#77

Bromobenzene

Concen: 18.477 ug/l

RT: 12.75 min Scan# 1779

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampleId :


Tot Ion:156 Resp: 46894

Ion Ratio Lower Upper

156 100

77 195.4 94.5 283.5




158 98.9 47.9 143.6



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VW

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 19.061 ug/l

RT: 12.80 min Scan# 1788

Delta R.T. 0.00 min

Lab File: VW015741.D


Acq: 26 Jun 2020 11:31

Tgt Ion: 91 Resp: 261151

Ion Ratio Lower Upper



91 100


120 22.8 11.2 33.5



Abundance Ion 91.00 (90.70 to 91.70): VW

Ion 120.00 (119.70 to 120.70): V









#79
2-Chlorotoluene
Concen: 19.502 ug/l
RT: 12.89 min Scan# 1802
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampled :


Tot Ion: 91 Resp: 150443
Ion Ratio Lower Upper
91 100
126 31.9 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 19.086 ug/l
RT: 12.94 min Scan# 1811
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:105 Resp: 184077
Ion Ratio Lower Upper
105 100
120 49.9 24.4 73.2









#81
trans-1,4-Dichloro-2-butene
Concen: 16.031 ug/l
RT: 12.51 min Scan# 1740
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 75 Resp: 8658
Ion Ratio Lower Upper
75 100
53 123.1 83.8 125.8
89 46.9 36.6 55.0



#82
4-Chlorotoluene
Concen: 19.175 ug/l
RT: 12.99 min Scan# 1819
Delta R.T. 0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion: 91 Resp: 156490
Ion Ratio Lower Upper
91 100
126 32.9 16.3 48.9









#83
tert-Butylbenzene
Concen: 19.214 ug/l
RT: 13.21 min Scan# 1854
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion:119 Resp: 156571
Ion Ratio Lower Upper
119 100
91 68.3 34.7 104.1
134 25.3 13.8 41.4



#84
1,2,4-Trimethylbenzene
Concen: 19.475 ug/l
RT: 13.25 min Scan# 1861
Delta R.T. -0.01 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:105 Resp: 187310
Ion Ratio Lower Upper
105 100
120 45.3 22.8 68.3









#85
 sec-Butylbenzene
 Concen: 19.377 ug/l
 RT: 13.38 min Scan# 1883
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampled :


Tot Ion:105 Resp: 225671
 Ion Ratio Lower Upper
 105 100
 134 19.6 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 19.361 ug/l
 RT: 13.50 min Scan# 1902
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion:119 Resp: 204869
 Ion Ratio Lower Upper
 119 100
 134 26.0 13.5 40.4
 91 24.5 12.3 36.8









#87
 1,3-Dichlorobenzene
 Concen: 19.119 ug/l
 RT: 13.50 min Scan# 1902
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampleId :


Tot Ion:146 Resp: 96327
 Ion Ratio Lower Upper
 146 100
 111 41.2 20.3 60.8
 148 64.7 31.6 94.7



#88
 1,4-Dichlorobenzene
 Concen: 19.095 ug/l
 RT: 13.58 min Scan# 1915
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion:146 Resp: 95281
 Ion Ratio Lower Upper
 146 100
 111 41.1 20.3 60.9
 148 64.5 33.7 101.1





#89

n-Butylbenzene

Concen: 19.499 ug/l

RT: 13.82 min Scan# 1955

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampleId :





Tot Ion: 91 Resp: 198061

Ion Ratio Lower Upper

91 100

92 53.3 27.5 82.4

134 25.0 13.0 38.9



#90

Hexachloroethane

Concen: 18.044 ug/l

RT: 14.10 min Scan# 2000

Delta R.T. 0.00 min

Lab File: VW015741.D




Acq: 26 Jun 2020 11:31


Tgt Ion: 117 Resp: 33977

Ion Ratio Lower Upper

117 100

201 73.6 37.1 111.5





#91

1,2-Dichlorobenzene

Concen: 19.280 ug/l

RT: 13.87 min Scan# 1963

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31

Instrument :

MSVOA_W

ClientSampleId :


Tot Ion:146 Resp: 83989

Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.7




148 64.2 32.2 96.6



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 17.621 ug/l

RT: 14.49 min Scan# 2064

Delta R.T. 0.00 min

Lab File: VW015741.D

Acq: 26 Jun 2020 11:31


Tgt Ion: 75 Resp: 5138

Ion Ratio Lower Upper

75 100

155 76.7 40.8 122.2



157 100.4 50.7 152.1




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 18.450 ug/l
 RT: 15.13 min Scan# 2170
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31




Instrument :
 MSVOA_W
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
180	100		
182	97.2	47.9	143.6
145	30.0	15.6	46.7



#94
 Hexachlorobutadiene
 Concen: 18.412 ug/l
 RT: 15.24 min Scan# 2187
 Delta R.T. 0.00 min
 Lab File: VW015741.D
 Acq: 26 Jun 2020 11:31

Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.2	31.0	93.0
227	65.1	31.9	95.9









#95
Naphthalene
Concen: 18.273 ug/l
RT: 15.37 min Scan# 2209
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Instrument :
MSVOA_W
ClientSampled :

Tot Ion:128 Resp: 86941
Ion Ratio Lower Upper
128 100
127 13.2 10.5 15.7
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 18.374 ug/l
RT: 15.56 min Scan# 2240
Delta R.T. 0.00 min
Lab File: VW015741.D
Acq: 26 Jun 2020 11:31

Tgt Ion:180 Resp: 45509
Ion Ratio Lower Upper
180 100
182 93.8 49.0 147.0
145 31.8 16.4 49.2

