


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062620\
 Data File : VW015737.D
 Acq On : 26 Jun 2020 09:20
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062320S.M
 Title : SW846 8260
 Last Update : Tue Jun 23 13:20:19 2020



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	19343	PASS
75	95	30	60	54.7	48418	PASS
95	95	100	100	100.0	88536	PASS
96	95	5	9	6.8	6008	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.1	68232	PASS
175	174	5	9	8.1	5503	PASS
176	174	95	101	99.7	68034	PASS
177	176	5	9	6.7	4574	PASS