

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062718\
 Data File : VW003582.D
 Acq On : 26 Jun 2018 21:57
 Operator : SY/AP
 Sample : J3632-02
 Misc : 7.03G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-105-4(65-67)

Quant Time: Jun 27 07:59:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	15938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	23576	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	21593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	8244	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	8127	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
35) Dibromofluoromethane	7.89	113	6779	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	10.33	98	17883	30.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.58%	
62) 4-Bromofluorobenzene	12.63	95	5426	26.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.80%	
Target Compounds						
16) Acetone	4.14	43	6666	184.11	ug/l	Qvalue 100
20) Methylene Chloride	4.92	84	2103	10.80	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062718\
Data File : VW003582.D
Acq On : 26 Jun 2018 21:57
Operator : SY/AP
Sample : J3632-02
Misc : 7.03G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-105-4(65-67)

Quant Time: Jun 27 07:59:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

