

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062718\
 Data File : VW003584.D
 Acq On : 26 Jun 2018 22:48
 Operator : SY/AP
 Sample : J3684-07
 Misc : 6.27G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-2-S

Quant Time: Jun 27 08:03:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	146361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	246875	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	204008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	72038	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.32	65	288	0.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.36%	
35) Dibromofluoromethane	7.87	113	123	0.08	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	0.16%	
50) Toluene-d8	10.34	98	1079	0.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.36%	
62) 4-Bromofluorobenzene	12.62	95	714	0.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.66%	
Target Compounds						
16) Acetone	4.14	43	23127	69.56	ug/l	98
17) Carbon Disulfide	4.37	76	4238	1.08	ug/l #	91
20) Methylene Chloride	4.92	84	8676	2.67	ug/l	96
25) 2-Butanone	7.18	43	3974	9.61	ug/l	92
40) Benzene	8.33	78	14655	2.23	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062718\
Data File : VW003584.D
Acq On : 26 Jun 2018 22:48
Operator : SY/AP
Sample : J3684-07
Misc : 6.27G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-2-S

Quant Time: Jun 27 08:03:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

