

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062718\
 Data File : VW003596.D
 Acq On : 27 Jun 2018 03:58
 Operator : SY/AP
 Sample : J3628-01RE
 Misc : 6.86G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 AOC11-01RE

Quant Time: Jun 27 08:21:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	107900	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	173752	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	137522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	46128	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	55976	47.89	ug/l	0.00
Spiked Amount			Recovery	=	95.78%	
35) Dibromofluoromethane	7.88	113	47723	44.20	ug/l	0.00
Spiked Amount			Recovery	=	88.40%	
50) Toluene-d8	10.33	98	128887	30.11	ug/l	0.00
Spiked Amount			Recovery	=	60.22%	
62) 4-Bromofluorobenzene	12.62	95	35152	23.20	ug/l	0.00
Spiked Amount			Recovery	=	46.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.90	84	13788	10.33	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062718\
Data File : VW003596.D
Acq On : 27 Jun 2018 03:58
Operator : SY/AP
Sample : J3628-01RE
Misc : 6.86G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
AOC11-01RE

Quant Time: Jun 27 08:21:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

