

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062718\
 Data File : VW003624.D
 Acq On : 27 Jun 2018 19:50
 Operator : SY/AP
 Sample : J3048-15
 Misc : 5.59G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-2-D

Quant Time: Jun 28 07:07:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	133355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	219622	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	187706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	68787	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	58609	40.58	ug/l	0.00
Spiked Amount			Recovery	=	81.16%	
35) Dibromofluoromethane	7.89	113	51366	37.64	ug/l	0.00
Spiked Amount			Recovery	=	75.28%	
50) Toluene-d8	10.33	98	172612	31.90	ug/l	0.00
Spiked Amount			Recovery	=	63.80%	
62) 4-Bromofluorobenzene	12.62	95	54616	28.52	ug/l	0.00
Spiked Amount			Recovery	=	57.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.15	43	21028	69.41	ug/l	90
17) Carbon Disulfide	4.38	76	5182	1.45	ug/l #	90
39) Methylcyclohexane	9.34	83	10048	4.09	ug/l	95
83) tert-Butylbenzene	13.21	119	5045	1.47	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062718\
Data File : VW003624.D
Acq On : 27 Jun 2018 19:50
Operator : SY/AP
Sample : J3048-15
Misc : 5.59G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-2-D

Quant Time: Jun 28 07:07:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

