

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062718\
 Data File : VW003627.D
 Acq On : 27 Jun 2018 21:07
 Operator : SY/AP
 Sample : J3709-03MSD
 Misc : 4.55G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EME-FW-TS004-01

Quant Time: Jun 28 07:13:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	30637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	45171	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	35305	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	10523	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	17778	53.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.14%	
35) Dibromofluoromethane	7.88	113	15378	54.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.58%	
50) Toluene-d8	10.33	98	33832	30.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	60.80%	
62) 4-Bromofluorobenzene	12.63	95	9015	22.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	45.78%	

Target Compounds						Qvalue
20) Methylene Chloride	4.90	84	3338	8.22	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062718\
Data File : VW003627.D
Acq On : 27 Jun 2018 21:07
Operator : SY/AP
Sample : J3709-03MSD
Misc : 4.55G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EME-FW-TS004-01

Quant Time: Jun 28 07:13:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

