

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062818\
 Data File : VW003631.D
 Acq On : 28 Jun 2018 11:54
 Operator : SY/AP
 Sample : VW0628SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0628SBL01

Quant Time: Jun 29 06:58:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	200870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	332887	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	303149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	147421	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	123270	56.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.32%	
35) Dibromofluoromethane	7.88	113	107549	51.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.98%	
50) Toluene-d8	10.33	98	409111	49.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.63	95	136802	47.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.26%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062818\
Data File : VW003631.D
Acq On : 28 Jun 2018 11:54
Operator : SY/AP
Sample : VW0628SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0628SBL01

Quant Time: Jun 29 06:58:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

