

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062818\
 Data File : VW003642.D
 Acq On : 28 Jun 2018 16:50
 Operator : SY/AP
 Sample : J3048-10RE
 Misc : 5.90G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-3-DRE

Quant Time: Jun 29 07:43:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	133993	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	206960	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	164155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	58875	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	80561	55.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.02%	
35) Dibromofluoromethane	7.88	113	72032	56.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.02%	
50) Toluene-d8	10.33	98	231199	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
62) 4-Bromofluorobenzene	12.62	95	70503	39.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.14%	
Target Compounds						
16) Acetone	4.15	43	19988	65.67	ug/l	95
20) Methylene Chloride	4.90	84	9063	3.60	ug/l	92
25) 2-Butanone	7.18	43	4024	10.63	ug/l #	84
86) p-Isopropyltoluene	13.51	119	11831	3.08	ug/l	93
95) Naphthalene	15.38	128	168925	82.66	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062818\
Data File : VW003642.D
Acq On : 28 Jun 2018 16:50
Operator : SY/AP
Sample : J3048-10RE
Misc : 5.90G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-3-DRE

Quant Time: Jun 29 07:43:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

