

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062818\
 Data File : VW003646.D
 Acq On : 28 Jun 2018 18:33
 Operator : SY/AP
 Sample : J3732-03
 Misc : 4.040G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-106-8(18-20)

Quant Time: Jun 29 08:04:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	140144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	231245	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	199411	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	81661	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	70030	46.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.26%	
35) Dibromofluoromethane	7.88	113	65387	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	
50) Toluene-d8	10.33	98	193499	33.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	67.92%	
62) 4-Bromofluorobenzene	12.62	95	71516	35.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.94%	
Target Compounds						
16) Acetone	4.15	43	42197	132.54	ug/l	95
17) Carbon Disulfide	4.37	76	31730	8.46	ug/l	100
30) Chloroform	7.68	83	9842	3.68	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062818\
Data File : VW003646.D
Acq On : 28 Jun 2018 18:33
Operator : SY/AP
Sample : J3732-03
Misc : 4.040G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-106-8(18-20)

Quant Time: Jun 29 08:04:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

