

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062818\
 Data File : VW003648.D
 Acq On : 28 Jun 2018 19:25
 Operator : SY/AP
 Sample : J3732-02RE
 Misc : 5.14G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-106-8(14-16)RE

Quant Time: Jun 29 08:09:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 05:18:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	141950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	231784	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	197637	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	79587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	71883	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
35) Dibromofluoromethane	7.89	113	66923	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
50) Toluene-d8	10.33	98	194851	34.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.24%	
62) 4-Bromofluorobenzene	12.63	95	63756	31.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.10%	
Target Compounds						
16) Acetone	4.14	43	36609	113.53	ug/l	98
17) Carbon Disulfide	4.37	76	24094	6.34	ug/l	96
20) Methylene Chloride	4.91	84	6741	1.35	ug/l	90
95) Naphthalene	15.38	128	5300	1.92	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062818\
Data File : VW003648.D
Acq On : 28 Jun 2018 19:25
Operator : SY/AP
Sample : J3732-02RE
Misc : 5.14G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-106-8(14-16)RE

Quant Time: Jun 29 08:09:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W061318S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 05:18:54 2018
Response via : Initial Calibration

