


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW062819\
Data File : VW011019.D
Acq On : 27 Jun 2019 13:59
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W062819S.M
Title : SW846 8260
Last Update : Fri Jun 28 01:55:46 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.7	17111	PASS
75	95	30	60	54.7	39509	PASS
95	95	100	100	100.0	72261	PASS
96	95	5	9	6.9	4981	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.5	54586	PASS
175	174	5	9	7.8	4277	PASS
176	174	95	101	97.3	53117	PASS
177	176	5	9	6.8	3613	PASS