

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062920\
 Data File : VW015765.D
 Acq On : 29 Jun 2020 16:21
 Operator : SY/VA
 Sample : L3131-05
 Misc : 5.51G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG2M0

Quant Time: Jun 30 03:49:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062920S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 30 03:03:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	343326	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	307815	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	143746	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	126915	22.77	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.08%
7) Chloroethane-d5	2.90	69	102841	22.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.36%
10) 1,1-Dichloroethene-d2	4.02	63	174089	17.59	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.36%
20) 2-Butanone-d5	7.09	46	47738	40.59	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.18%
24) Chloroform-d	7.65	84	215849	22.55	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	8.31	65	120567	22.15	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.60%
29) Benzene-d6	8.28	84	439373	23.94	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.76%
33) 1,2-Dichloropropane-d6	9.28	67	126461	23.27	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.08%
37) Toluene-d8	10.33	98	408618	23.93	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.72%
38) trans-1,3-Dichloropropene-	10.58	79	47986	19.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.76%
39) 2-Hexanone-d5	10.93	63	35408	39.99	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.98%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	86208	20.69	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.76%
61) 1,2-Dichlorobenzene-d4	13.85	152	134346	24.10	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.40%

Target Compounds

					Ovalue
13) Acetone	4.13	43	22557	22.885	ug/L 100
14) Carbon disulfide	4.38	76	102480	8.265	ug/L 99
16) Methylene chloride	4.92	84	8161	1.620	ug/L 95
21) 2-Butanone	7.18	43	7052	5.787	ug/L 95
44) Toluene	10.39	91	29327	1.597	ug/L 100
53) Ethylbenzene	11.73	91	890362	42.877	ug/L 100
54) m,p-Xylene	11.84	106	1733103	223.294	ug/L 99
55) o-xylene	12.17	106	674053	90.869	ug/L 97
57) Isopropylbenzene	12.47	105	22236	1.085	ug/L 97
65) 1,2-Dichlorobenzene	13.87	146	4553	0.604	ug/L # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062920\
Data File : VW015765.D
Acq On : 29 Jun 2020 16:21
Operator : SY/VA
Sample : L3131-05
Misc : 5.51G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG2M0

Quant Time: Jun 30 03:49:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062920S.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 30 03:03:23 2020
Response via : Initial Calibration

