

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062920\
 Data File : VW015768.D
 Acq On : 29 Jun 2020 17:57
 Operator : SY/VA
 Sample : L3131-07
 Misc : 5.12G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG2M1

Quant Time: Jun 30 04:35:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062920S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 30 03:03:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	375907	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	336586	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	146419	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	129580	21.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.92%
7) Chloroethane-d5	2.89	69	104809	21.03	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.12%
10) 1,1-Dichloroethene-d2	4.01	63	174092	16.06	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.24%
20) 2-Butanone-d5	7.09	46	41772	32.44	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.88%
24) Chloroform-d	7.65	84	213696	20.39	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.56%
26) 1,2-Dichloroethane-d4	8.31	65	113789	19.09	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.36%
29) Benzene-d6	8.27	84	430549	21.45	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.80%
33) 1,2-Dichloropropane-d6	9.27	67	123732	20.82	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.28%
37) Toluene-d8	10.32	98	394197	21.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.44%
38) trans-1,3-Dichloropropene-	10.58	79	47593	18.08	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.32%
39) 2-Hexanone-d5	10.93	63	30144	31.13	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.26%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	75037	16.47	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	65.88%
61) 1,2-Dichlorobenzene-d4	13.85	152	119646	21.07	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.28%

Target Compounds

					Ovalue
13) Acetone	4.14	43	36281	33.619	ug/L 98
14) Carbon disulfide	4.38	76	267861	19.729	ug/L 100
16) Methylene chloride	4.92	84	11983	2.173	ug/L 93
21) 2-Butanone	7.18	43	14564	10.916	ug/L 91
53) Ethylbenzene	11.73	91	59516	2.621	ug/L 97
54) m,p-Xylene	11.84	106	3946	0.465	ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062920\
Data File : VW015768.D
Acq On : 29 Jun 2020 17:57
Operator : SY/VA
Sample : L3131-07
Misc : 5.12G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG2M1

Quant Time: Jun 30 04:35:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062920S.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 30 03:03:23 2020
Response via : Initial Calibration

