

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070124\
 Data File : VW029429.D
 Acq On : 01 Jul 2024 16:05
 Operator : SY/MD
 Sample : P3100-05
 Misc : 5.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jul 02 08:11:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jun 29 02:05:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	259688	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	243851	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	112153	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	59232	16.479	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	65.920%		
7) Chloroethane-d5	2.899	69	51516	16.972	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	67.880%		
11) 1,1-Dichloroethene-d2	4.021	65	29566	17.036	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	68.160%		
21) 2-Butanone-d5	7.081	46	53960	50.178	ug/L	0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	100.360%		
24) Chloroform-d	7.648	84	155094	20.021	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.080%		
26) 1,2-Dichloroethane-d4	8.307	65	96102	21.607	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	86.440%		
32) Benzene-d6	8.276	84	298038	20.122	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	80.480%		
36) 1,2-Dichloropropane-d6	9.276	67	101035	21.459	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	85.840%		
41) Toluene-d8	10.325	98	261166	19.930	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	79.720%		
43) trans-1,3-Dichloroprop...	10.575	79	38070	20.065	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	80.280%		
47) 2-Hexanone-d5	10.922	63	35935	48.687	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	97.380%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	84972	22.676	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	90.720%		
66) 1,2-Dichlorobenzene-d4	13.848	152	89736	23.051	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	92.200%		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	11302	2.475	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070124\
Data File : VW029429.D
Acq On : 01 Jul 2024 16:05
Operator : SY/MD
Sample : P3100-05
Misc : 5.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jul 02 08:11:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jun 29 02:05:50 2024
Response via : Initial Calibration

