

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070124\
 Data File : VW029432.D
 Acq On : 01 Jul 2024 17:16
 Operator : SY/MD
 Sample : P3125-01
 Misc : 5.64g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4CZ2

Quant Time: Jul 02 08:12:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jun 29 02:05:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	440661	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	330718	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	98668	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	113118	18.546	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.200%
7) Chloroethane-d5	2.899	69	95396	18.521	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.080%
11) 1,1-Dichloroethene-d2	4.015	65	54470	18.496	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.000%
21) 2-Butanone-d5	7.075	46	73692	40.384	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.760%
24) Chloroform-d	7.648	84	257339	19.577	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.320%
26) 1,2-Dichloroethane-d4	8.307	65	150662	19.962	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	79.840%
32) Benzene-d6	8.276	84	470994	23.447	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.800%
36) 1,2-Dichloropropane-d6	9.270	67	161810	25.340	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.360%
41) Toluene-d8	10.319	98	387216	21.788	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.160%
43) trans-1,3-Dichloroprop...	10.575	79	48907	19.006	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.040%
47) 2-Hexanone-d5	10.922	63	38969	38.930	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.860%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	104717	20.605	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	13.849	152	74359	21.712	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070124\
Data File : VW029432.D
Acq On : 01 Jul 2024 17:16
Operator : SY/MD
Sample : P3125-01
Misc : 5.64g/10mL/MSVOA_W/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4CZ2

Quant Time: Jul 02 08:12:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061824SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jun 29 02:05:50 2024
Response via : Initial Calibration

