

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070318\
 Data File : VW003695.D
 Acq On : 03 Jul 2018 13:40
 Operator : SY/AP
 Sample : J3252-05
 Misc : 6.72G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TR-04-062918

Quant Time: Jul 03 23:56:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	192816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	323940	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	283559	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	126650	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	109378	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
35) Dibromofluoromethane	7.88	113	92530	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
50) Toluene-d8	10.33	98	300179	36.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.38%	
62) 4-Bromofluorobenzene	12.63	95	101913	34.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.40%	
Target Compounds						
16) Acetone	4.14	43	13245	24.138	ug/l	Qvalue 100
20) Methylene Chloride	4.92	84	9786	1.413	ug/l #	84
84) 1,2,4-Trimethylbenzene	13.26	105	7683	1.024	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070318\
Data File : VW003695.D
Acq On : 03 Jul 2018 13:40
Operator : SY/AP
Sample : J3252-05
Misc : 6.72G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TR-04-062918

Quant Time: Jul 03 23:56:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

