

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070318\
 Data File : VW003698.D
 Acq On : 03 Jul 2018 14:58
 Operator : SY/AP
 Sample : J3245-19
 Misc : 6.43G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-01-062918-B

Quant Time: Jul 04 00:01:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	170373	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	281391	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	237741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	93790	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	102669	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	7.88	113	86448	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
50) Toluene-d8	10.33	98	266866	37.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.08%	
62) 4-Bromofluorobenzene	12.62	95	83301	32.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.36%	
Target Compounds						
20) Methylene Chloride	4.93	84	9711	2.027	ug/l	91
68) m/p-Xylenes	11.84	106	6391	1.919	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070318\
Data File : VW003698.D
Acq On : 03 Jul 2018 14:58
Operator : SY/AP
Sample : J3245-19
Misc : 6.43G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-01-062918-B

Quant Time: Jul 04 00:01:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

