

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW070318\
 Data File : VW003688.D
 Acq On : 03 Jul 2018 10:16
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 13:22:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.239	0.329	-37.7#	150#	0.00
3 P	Chloromethane	0.398	0.438	-10.1	118	0.00
4 C	Vinyl Chloride	0.543	0.547	-0.7#	109	0.00
5 T	Bromomethane	0.356	0.373	-4.8	115	0.00
6 T	Chloroethane	0.327	0.327	0.0	111	0.00
7 T	Trichlorofluoromethane	0.289	0.304	-5.2	119	0.00
8 T	Diethyl Ether	0.267	0.247	7.5	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.480	0.478	0.4	112	0.00
10 T	Methyl Iodide	0.705	0.710	-0.7	110	0.00
11 T	Tert butyl alcohol	0.037	0.032	13.5	97	-0.02
12 CM	1,1-Dichloroethene	0.479	0.473	1.3#	110	0.00
13 T	Acrolein	0.018	0.022	-22.2#	215#	0.00
14 T	Allyl chloride	0.819	0.840	-2.6	113	0.00
15 T	Acrylonitrile	0.124	0.113	8.9	102	0.00
16 T	Acetone	0.142	0.143	-0.7	118	0.00
17 T	Carbon Disulfide	1.437	1.532	-6.6	113	0.00
18 T	Methyl Acetate	0.334	0.259	22.5#	102	0.00
19 T	Methyl tert-butyl Ether	0.935	0.906	3.1	105	0.00
20 T	Methylene Chloride	0.645	0.536	16.9	111	0.00
21 T	trans-1,2-Dichloroethene	0.524	0.518	1.1	108	0.00
22 T	Diisopropyl ether	1.678	1.640	2.3	105	0.00
23 T	Vinyl Acetate	1.026	1.029	-0.3	104	0.00
24 P	1,1-Dichloroethane	1.001	0.963	3.8	107	0.00
25 T	2-Butanone	0.182	0.170	6.6	104	0.00
26 T	2,2-Dichloropropane	0.614	0.598	2.6	113	0.00
27 T	cis-1,2-Dichloroethene	0.571	0.563	1.4	109	0.00
28 T	Bromochloromethane	0.440	0.400	9.1	100	0.00
29 T	Tetrahydrofuran	0.109	0.100	8.3	100	0.00
30 C	Chloroform	1.012	0.966	4.5#	107	0.00
31 T	Cyclohexane	0.911	0.913	-0.2	114	0.00
32 T	1,1,1-Trichloroethane	0.788	0.787	0.1	112	0.00
33 S	1,2-Dichloroethane-d4	0.596	0.580	2.7	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.326	0.332	-1.8	107	0.00
36 T	1,1-Dichloropropene	0.502	0.526	-4.8	110	0.00
37 T	Ethyl Acetate	0.244	0.239	2.0	105	0.00
38 T	Carbon Tetrachloride	0.477	0.500	-4.8	111	0.00
39 T	Methylcyclohexane	0.578	0.629	-8.8	111	0.00
40 TM	Benzene	1.411	1.436	-1.8	108	0.00
41 T	Methacrylonitrile	0.139	0.136	2.2	103	0.00
42 TM	1,2-Dichloroethane	0.444	0.440	0.9	106	0.00
43 T	Isopropyl Acetate	0.460	0.441	4.1	102	0.00
44 TM	Trichloroethene	0.365	0.371	-1.6	110	0.00
45 C	1,2-Dichloropropane	0.366	0.362	1.1#	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW070318\
 Data File : VW003688.D
 Acq On : 03 Jul 2018 10:16
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 13:22:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.190	0.187	1.6	105	0.00
47 T	Bromodichloromethane	0.470	0.468	0.4	106	0.00
48 T	Methyl methacrylate	0.218	0.215	1.4	101	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	100	0.00
50 S	Toluene-d8	1.280	1.352	-5.6	110	0.00
51 T	4-Methyl-2-Pentanone	0.242	0.234	3.3	100	0.00
52 CM	Toluene	0.880	0.917	-4.2#	108	0.00
53 T	t-1,3-Dichloropropene	0.487	0.499	-2.5	106	0.00
54 T	cis-1,3-Dichloropropene	0.553	0.562	-1.6	105	0.00
55 T	1,1,2-Trichloroethane	0.268	0.260	3.0	104	0.00
56 T	Ethyl methacrylate	0.336	0.340	-1.2	102	0.00
57 T	1,3-Dichloropropane	0.465	0.457	1.7	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.135	5.6	100	0.00
59 T	2-Hexanone	0.173	0.177	-2.3	105	0.00
60 T	Dibromochloromethane	0.313	0.310	1.0	105	0.00
61 T	1,2-Dibromoethane	0.256	0.250	2.3	104	0.00
62 S	4-Bromofluorobenzene	0.460	0.482	-4.8	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.338	0.342	-1.2	108	0.00
65 PM	Chlorobenzene	1.051	1.066	-1.4	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.373	-1.6	106	0.00
67 C	Ethyl Benzene	1.842	1.943	-5.5#	108	0.00
68 T	m/p-Xylenes	0.700	0.738	-5.4	108	0.00
69 T	o-Xylene	0.647	0.684	-5.7	108	0.00
70 T	Styrene	1.088	1.151	-5.8	105	0.00
71 P	Bromoform	0.204	0.202	1.0	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.401	3.661	-7.6	110	0.00
74 T	N-amyl acetate	0.898	0.873	2.8	99	0.00
75 P	1,1,2,2-Tetrachloroethane	0.665	0.636	4.4	102	0.00
76 T	1,2,3-Trichloropropane	0.450	0.453	-0.7	107	0.00
77 T	Bromobenzene	0.816	0.830	-1.7	105	0.00
78 T	n-propylbenzene	4.202	4.481	-6.6	108	0.00
79 T	2-Chlorotoluene	2.391	2.497	-4.4	107	0.00
80 T	1,3,5-Trimethylbenzene	2.877	3.078	-7.0	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.212	0.211	0.5	103	0.00
82 T	4-Chlorotoluene	2.538	2.628	-3.5	107	0.00
83 T	tert-Butylbenzene	2.427	2.617	-7.8	109	0.00
84 T	1,2,4-Trimethylbenzene	2.962	3.146	-6.2	107	0.00
85 T	sec-Butylbenzene	3.612	3.856	-6.8	108	0.00
86 T	p-Isopropyltoluene	3.159	3.396	-7.5	108	0.00
87 T	1,3-Dichlorobenzene	1.657	1.667	-0.6	106	0.00
88 T	1,4-Dichlorobenzene	1.655	1.653	0.1	105	0.00
89 T	n-Butylbenzene	3.123	3.375	-8.1	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW070318\
Data File : VW003688.D
Acq On : 03 Jul 2018 10:16
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Jul 03 13:22:53 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.587	0.606	-3.2	107	0.00
91 T	1,2-Dichlorobenzene	1.481	1.478	0.2	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.116	0.108	6.9	99	0.00
93 T	1,2,4-Trichlorobenzene	1.044	1.054	-1.0	105	0.00
94 T	Hexachlorobutadiene	0.603	0.626	-3.8	109	0.00
95 T	Naphthalene	1.894	1.888	0.3	100	0.00
96 T	1,2,3-Trichlorobenzene	0.930	0.909	2.3	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6