

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070518\
 Data File : VW003716.D
 Acq On : 05 Jul 2018 13:07
 Operator : SY/AP
 Sample : J3840-03
 Misc : 7.23G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-6-S

Quant Time: Jul 06 03:03:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	210119	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	347306	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	308153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	143603	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	112791	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
35) Dibromofluoromethane	7.88	113	102317	45.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.34%	
50) Toluene-d8	10.33	98	340950	38.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.68%	
62) 4-Bromofluorobenzene	12.63	95	117375	36.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.48%	
Target Compounds						
16) Acetone	4.14	43	22504	37.634	ug/l	98
20) Methylene Chloride	4.91	84	23322	7.334	ug/l	94
64) Tetrachloroethene	10.86	164	4879	2.343	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070518\
Data File : VW003716.D
Acq On : 05 Jul 2018 13:07
Operator : SY/AP
Sample : J3840-03
Misc : 7.23G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-6-S

Quant Time: Jul 06 03:03:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

