

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070518\
 Data File : VW003719.D
 Acq On : 05 Jul 2018 14:25
 Operator : SY/AP
 Sample : J3840-17
 Misc : 6.53G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-4-D

Quant Time: Jul 06 03:51:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	157593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	255350	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	188210	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	51739	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	86850	46.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.46%	
35) Dibromofluoromethane	7.88	113	77241	46.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.76%	
50) Toluene-d8	10.33	98	218196	33.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.74%	
62) 4-Bromofluorobenzene	12.62	95	48707	20.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	41.48%	
Target Compounds						
16) Acetone	4.14	43	4772	10.640	ug/l #	80
20) Methylene Chloride	4.91	84	12825	4.423	ug/l	97
64) Tetrachloroethene	10.87	164	3964	3.116	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070518\
Data File : VW003719.D
Acq On : 05 Jul 2018 14:25
Operator : SY/AP
Sample : J3840-17
Misc : 6.53G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-4-D

Quant Time: Jul 06 03:51:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

