

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070518\
 Data File : VW003721.D
 Acq On : 05 Jul 2018 15:17
 Operator : SY/AP
 Sample : J3814-01
 Misc : 3.79G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 8782-AA

Quant Time: Jul 06 03:58:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	125790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	210561	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	199410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	88376	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	76477	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
35) Dibromofluoromethane	7.88	113	65748	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
50) Toluene-d8	10.33	98	197000	36.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.08%	
62) 4-Bromofluorobenzene	12.62	95	68029	35.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.24%	
Target Compounds						
20) Methylene Chloride	4.90	84	10897	4.939	ug/l	95
30) Chloroform	7.68	83	3480	1.367	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070518\
Data File : VW003721.D
Acq On : 05 Jul 2018 15:17
Operator : SY/AP
Sample : J3814-01
Misc : 3.79G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
8782-AA

Quant Time: Jul 06 03:58:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

