

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070518\
 Data File : VW003727.D
 Acq On : 05 Jul 2018 17:52
 Operator : SY/AP
 Sample : J3813-01RE
 Misc : 4.28G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 I2-60-SRI3RE

Quant Time: Jul 06 12:20:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 03 08:55:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	17407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	25704	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	24560	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	10386	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	10365	49.95	ug/l	0.00
Spiked Amount			Recovery	=	99.90%	
35) Dibromofluoromethane	7.89	113	8719	52.01	ug/l	0.00
Spiked Amount			Recovery	=	104.02%	
50) Toluene-d8	10.33	98	22123	33.62	ug/l	0.00
Spiked Amount			Recovery	=	67.24%	
62) 4-Bromofluorobenzene	12.63	95	7712	32.62	ug/l	0.00
Spiked Amount			Recovery	=	65.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.91	84	1485	4.81	ug/l	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070518\
Data File : VW003727.D
Acq On : 05 Jul 2018 17:52
Operator : SY/AP
Sample : J3813-01RE
Misc : 4.28G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
I2-60-SRI3RE

Quant Time: Jul 06 12:20:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070218S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 03 08:55:34 2018
Response via : Initial Calibration

