

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070722\
 Data File : VW023581.D
 Acq On : 07 Jul 2022 18:21
 Operator : SY/VA
 Sample : N3626-05
 Misc : 5.61g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0ZC1

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/08/2022
 Supervised By : Mahesh Dadoda 07/08/2022

Quant Time: Jul 08 01:48:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 08 01:46:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	342205	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	308458	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	112442	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	132998	21.470	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	85.880%		
7) Chloroethane-d5	2.879	69	83390	23.538	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	94.160%		
11) 1,1-Dichloroethene-d2	4.013	63	103347	14.798	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	59.200%		
21) 2-Butanone-d5	7.080	46	67909m	52.791	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	105.580%		
24) Chloroform-d	7.653	84	219369	22.506	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	90.040%		
26) 1,2-Dichloroethane-d4	8.305	65	125892	22.851	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	91.400%		
32) Benzene-d6	8.274	84	402820	23.602	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	94.400%		
36) 1,2-Dichloropropane-d6	9.274	67	122775	24.571	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	98.280%		
41) Toluene-d8	10.323	98	359586	21.737	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	86.960%		
43) trans-1,3-Dichloroprop...	10.579	79	44366	19.891	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	79.560%		
47) 2-Hexanone-d5	10.926	63	49587	50.741	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	101.480%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	124875	25.191	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	100.760%		
66) 1,2-Dichlorobenzene-d4	13.853	152	95196	23.466	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	93.880%		
Target Compounds					Qvalue	
13) Acetone	4.135	43	4659m	4.914	ug/L	
16) Methylene chloride	4.921	84	37489	6.460	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070722\
Data File : VW023581.D
Acq On : 07 Jul 2022 18:21
Operator : SY/VA
Sample : N3626-05
Misc : 5.61g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0ZC1

Quant Time: Jul 08 01:48:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 08 01:46:16 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/08/2022
Supervised By :Mahesh Dadoda 07/08/2022

