

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
 Data File : VW003780.D
 Acq On : 09 Jul 2018 16:26
 Operator : SY/AP
 Sample : J3874-02RE
 Misc : 5.01G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 2018-NYSEG-MCMaster-PH4-5(3-

Quant Time: Jul 09 17:33:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	73333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	120195	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	100721	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	33261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	155430	162.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	324.16%	
35) Dibromofluoromethane	7.88	113	138128	164.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	328.40%	
50) Toluene-d8	10.33	98	408464	126.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	253.88%	
62) 4-Bromofluorobenzene	12.62	95	112836	96.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	193.82%	
Target Compounds						
20) Methylene Chloride	4.90	84	14123	14.286	ug/l	96
44) Trichloroethene	9.09	130	2638	2.870	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003780.D
Acq On : 09 Jul 2018 16:26
Operator : SY/AP
Sample : J3874-02RE
Misc : 5.01G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-MCMaster-PH4-5(3-

Quant Time: Jul 09 17:33:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

