

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
 Data File : VW003781.D
 Acq On : 09 Jul 2018 16:52
 Operator : SY/AP
 Sample : J3864-01
 Misc : 5.48G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-108-6A(3-8)

Quant Time: Jul 09 17:36:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	72504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	114761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	99421	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	34072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	167821	177.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	354.00%	
35) Dibromofluoromethane	7.88	113	145113	180.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	361.34%	
50) Toluene-d8	10.33	98	440219	143.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	286.56%	
62) 4-Bromofluorobenzene	12.62	95	120982	108.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	217.66%	
Target Compounds						
16) Acetone	4.12	43	2536	15.416	ug/l	95
20) Methylene Chloride	4.90	84	11602	11.057	ug/l	96
30) Chloroform	7.68	83	4021	2.503	ug/l	98
52) Toluene	10.40	92	2833	1.312	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003781.D
Acq On : 09 Jul 2018 16:52
Operator : SY/AP
Sample : J3864-01
Misc : 5.48G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-108-6A(3-8)

Quant Time: Jul 09 17:36:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

