

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
 Data File : VW003784.D
 Acq On : 09 Jul 2018 18:10
 Operator : SY/AP
 Sample : J3891-01
 Misc : 4.15G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 212-72-11940-R4-VNJ-242

Quant Time: Jul 09 18:39:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	61998	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	98550	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	88317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	28434	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	145368	179.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	358.60%	
35) Dibromofluoromethane	7.88	113	126005	182.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	365.38%	
50) Toluene-d8	10.33	98	360182	136.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	273.04%	
62) 4-Bromofluorobenzene	12.62	95	101910	106.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	213.50%	
Target Compounds						
20) Methylene Chloride	4.91	84	18094	24.126	ug/l	96
30) Chloroform	7.68	83	4018	2.926	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003784.D
Acq On : 09 Jul 2018 18:10
Operator : SY/AP
Sample : J3891-01
Misc : 4.15G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
212-72-11940-R4-VNJ-242

Quant Time: Jul 09 18:39:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

