

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
 Data File : VW003786.D
 Acq On : 09 Jul 2018 19:02
 Operator : SY/AP
 Sample : J3893-01
 Misc : 7.06G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BS-S-7-SOIL

Quant Time: Jul 10 04:16:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	69825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	113975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	108804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	45828	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	159869	175.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	350.16%	
35) Dibromofluoromethane	7.88	113	138292	173.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	346.72%	
50) Toluene-d8	10.33	98	429952	140.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	281.82%	
62) 4-Bromofluorobenzene	12.62	95	143247	129.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	259.50%	
Target Compounds						
16) Acetone	4.13	43	18338	115.753	ug/l	96
20) Methylene Chloride	4.91	84	11327	11.275	ug/l	91
30) Chloroform	7.67	83	3702	2.393	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003786.D
Acq On : 09 Jul 2018 19:02
Operator : SY/AP
Sample : J3893-01
Misc : 7.06G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BS-S-7-SOIL

Quant Time: Jul 10 04:16:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

