

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
 Data File : VW003788.D
 Acq On : 09 Jul 2018 19:54
 Operator : SY/AP
 Sample : J3897-01
 Misc : 4.79G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 STOCK-PILE-A

Quant Time: Jul 10 04:18:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	124957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	209499	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	171410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	51067	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	84583	51.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.52%	
35) Dibromofluoromethane	7.88	113	72151	49.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.42%	
50) Toluene-d8	10.33	98	200500	35.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	71.50%	
62) 4-Bromofluorobenzene	12.63	95	51132	25.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	50.40%	
Target Compounds						
20) Methylene Chloride	4.91	84	8752	2.137	ug/l	90
30) Chloroform	7.67	83	3426	1.238	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003788.D
Acq On : 09 Jul 2018 19:54
Operator : SY/AP
Sample : J3897-01
Misc : 4.79G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
STOCK-PILE-A

Quant Time: Jul 10 04:18:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

