

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
 Data File : VW003791.D
 Acq On : 09 Jul 2018 21:12
 Operator : SY/AP
 Sample : J3886-02
 Misc : 6.01G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-108-6A(60-62)

Quant Time: Jul 10 04:20:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	74524	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	137910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	114514	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	27358	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	59252	60.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.60%	
35) Dibromofluoromethane	7.88	113	52284	54.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.34%	
50) Toluene-d8	10.33	98	125313	33.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.88%	
62) 4-Bromofluorobenzene	12.62	95	29236	21.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	43.78%	
Target Compounds						
16) Acetone	4.12	43	19534	115.528	ug/l	96
17) Carbon Disulfide	4.37	76	3617	1.576	ug/l #	92
20) Methylene Chloride	4.90	84	5838	2.960	ug/l	84
30) Chloroform	7.68	83	1936	1.173	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003791.D
Acq On : 09 Jul 2018 21:12
Operator : SY/AP
Sample : J3886-02
Misc : 6.01G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-108-6A(60-62)

Quant Time: Jul 10 04:20:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

