


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW070918\
Data File : VW003767.D
Acq On : 09 Jul 2018 09:21
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Title : SW846 8260
Last Update : Sat Jul 07 05:07:25 2018



AutoFind: Scans 1809, 1810, 1811; Background Corrected with Scan 1801

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	20530	PASS
75	95	30	60	52.7	49802	PASS
95	95	100	100	100.0	94536	PASS
96	95	5	9	6.7	6295	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.2	73925	PASS
175	174	5	9	7.6	5651	PASS
176	174	95	101	97.6	72184	PASS
177	176	5	9	7.0	5056	PASS