

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070921\
 Data File : VW019397.D
 Acq On : 09 Jul 2021 15:14
 Operator : SY/VA
 Sample : M2959-01
 Misc : 4.20G/10ML/MSVOA_W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG7F7

Quant Time: Jul 10 01:01:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jul 10 00:58:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	578198	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	530856	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	257531	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	153806	18.774	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	75.080%		
7) Chloroethane-d5	2.898	69	126805	21.706	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	86.840%		
11) 1,1-Dichloroethene-d2	4.025	63	246967	15.624	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	62.480%		
21) 2-Butanone-d5	7.080	46	103675	43.602	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	87.200%		
24) Chloroform-d	7.653	84	357447	22.374	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	89.480%		
26) 1,2-Dichloroethane-d4	8.311	65	202650	23.114	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	92.440%		
32) Benzene-d6	8.275	84	746004	23.608	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	94.440%		
36) 1,2-Dichloropropane-d6	9.275	67	233661	23.724	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	94.880%		
41) Toluene-d8	10.323	98	660238	23.174	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	92.680%		
43) trans-1,3-Dichloroprop...	10.579	79	94223	22.339	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	89.360%		
47) 2-Hexanone-d5	10.927	63	81512	44.609	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	89.220%		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	183560	23.065	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	92.240%		
66) 1,2-Dichlorobenzene-d4	13.853	152	221353	24.361	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	97.440%		
Target Compounds						Qvalue
16) Methylene chloride	4.928	84	39497	3.427	ug/L	91
46) Tetrachloroethene	10.866	164	5488	0.849	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW070921\
Data File : VW019397.D
Acq On : 09 Jul 2021 15:14
Operator : SY/VA
Sample : M2959-01
Misc : 4.20G/10ML/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG7F7

Quant Time: Jul 10 01:01:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jul 10 00:58:28 2021
Response via : Initial Calibration

