

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071018\
 Data File : VW003825.D
 Acq On : 10 Jul 2018 15:14
 Operator : SY/AP
 Sample : J3887-07RE
 Misc : 6.31G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-8-DRE

Quant Time: Jul 10 15:33:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	135156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	230923	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	218004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	102768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	72095	40.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.58%	
35) Dibromofluoromethane	7.88	113	63239	39.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.26%	
50) Toluene-d8	10.33	98	197138	31.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.78%	
62) 4-Bromofluorobenzene	12.62	95	69023	30.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.72%	
Target Compounds						
16) Acetone	4.14	43	12162	39.66	ug/l	98
20) Methylene Chloride	4.91	84	16896	7.59	ug/l	96
30) Chloroform	7.68	83	3925	1.31	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071018\
Data File : VW003825.D
Acq On : 10 Jul 2018 15:14
Operator : SY/AP
Sample : J3887-07RE
Misc : 6.31G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-8-DRE

Quant Time: Jul 10 15:33:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

