

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071018\
 Data File : VW003831.D
 Acq On : 10 Jul 2018 17:50
 Operator : SY/AP
 Sample : J3877-01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 NS-DRUM-2-STONE

Quant Time: Jul 11 05:10:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	116364	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	223979	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	183503	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	56756	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	78062	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
35) Dibromofluoromethane	7.88	113	62797	40.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.12%	
50) Toluene-d8	10.33	98	185449	30.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.86%	
62) 4-Bromofluorobenzene	12.62	95	54195	24.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.96%	
Target Compounds						
16) Acetone	4.14	43	15731	59.58	ug/l	99
20) Methylene Chloride	4.91	84	26102	17.43	ug/l	98
25) 2-Butanone	7.19	43	4090	9.89	ug/l	94
48) Methyl methacrylate	9.44	41	5606	5.37	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071018\
Data File : VW003831.D
Acq On : 10 Jul 2018 17:50
Operator : SY/AP
Sample : J3877-01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
NS-DRUM-2-STONE

Quant Time: Jul 11 05:10:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

